American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37623	05/30/2006	H&A

Method: 8260B, Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 060206

Our Lab I.D.			Method Blank	**37623.33**	**37623.34**		
Client Sample I.D.				EB-053006-01	TB-053006		
Date Sampled				05/30/2006	05/30/2006		
Date Prepared			06/02/2006	06/02/2006	06/02/2006		
Preparation Method			5030B	5030B	5030B		
Date Analyzed			06/02/2006	06/02/2006	06/02/2006		
Matrix			Aqueous	Aqueous	Aqueous		
Units			ug/L	ug/L	ug/L		
Dilution Factor			1	1	1		
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**		
Vinyl Acetate	0.5	5.0	ND	ND	ND		
Vinyl chloride (Chloroethene)	0.5	3.0	ND	ND	ND		
o-Xylene	0.5	1.0	ND	ND	ND		
m,p-Xylenes	1.0	2.0	ND	ND	ND		
Our Lab I.D.			Method Blank	**37623.33**	**37623.34**		
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**		
Bromofluorobenzene	75-125		90	92	91		
Dibromofluoromethane	75-125		119	121	122		
Toluene-d8	75-125		88	92	90		

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page 17
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37623	05/30/2006	H&A

Analytes			Antimony	Arsenic	Lead	
Methods of Analyses			6010BSCAN	6010BSCAN	6010BSCAN	
Date Prepared			06/01/2006	06/01/2006	06/01/2006	
Date Analyzed			06/01/2006	06/01/2006	06/01/2006	
Matrix			Aqueous	Aqueous	Aqueous	
QC Batch Number			060106	060106	060106	
Units			mg/L	mg/L	mg/L	
Method Detection Limit			0.05	0.05	0.03	
Practical Quantitation Limit			0.10	0.10	0.10	
Dilution Factor			1	1	1	
Lab ID	**Sample ID**	**Sampled**	**Results**	**Results**	**Results**	
37623.33	EB-053006-01	05/30/2006	ND	ND	ND	
N/A	Method Blank	05/30/2006	ND	ND	ND	

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **18**

Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37623	05/30/2006	H&A

Method: 8082, Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 060106

Our Lab I.D.			Method Blank	**37623.33**			
Client Sample I.D.				EB-053006-01			
Date Sampled				05/30/2006			
Date Prepared			06/01/2006	06/01/2006			
Preparation Method			3510C	3510C			
Date Analyzed			06/02/2006	06/02/2006			
Matrix			Aqueous	Aqueous			
Units			ug/L	ug/L			
Dilution Factor			1	1			
Analytes	**MDL**	**PQL**	**Results**	**Results**			
Aroclor-1016 (PCB-1016)	1.0	5.0	ND	ND			
Aroclor-1221 (PCB-1221)	2.0	10.0	ND	ND			
Aroclor-1232 (PCB-1232)	1.0	5.0	ND	ND			
Aroclor-1242 (PCB-1242)	1.0	5.0	ND	ND			
Aroclor-1248 (PCB-1248)	1.0	5.0	ND	ND			
Aroclor-1254 (PCB-1254)	1.0	5.0	ND	ND			
Aroclor-1260 (PCB-1260)	1.0	5.0	ND	ND			
Aroclor-1262 (PCB-1262)	1.0	5.0	ND	ND			
Aroclor-1268 (PCB-1268)	1.0	5.0	ND	ND			
Our Lab I.D.			Method Blank	**37623.33**			
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**			
Tetrachloro-m-xylene	30-150		50	54			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **19**
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37623	05/30/2006	H&A

Method: 6010BSCAN, 3 Metals(As,Pb,Sb)

QUALITY CONTROL REPORT

QC Batch No: 060106 Sample Spiked: 37640.02 QC Prepared: 06/01/2006 QC Analyzed: 06/01/2006 Units: mg/L

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Antimony	0.022	1.00	0.89	87	1.00	0.91	89	2.3	80-120	<15
Arsenic	0.005	1.00	0.94	93	1.00	0.91	90	3.3	80-120	<15
Lead	ND	1.00	0.83	83	1.00	0.83	83	<1	80-120	<15

QC Batch No: 060106 Sample Spiked: 37640.02 QC Prepared: 06/01/2006 QC Analyzed: 06/01/2006 Units: mg/L

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Antimony	1.00	0.82	82	80-120						
Arsenic	1.00	0.93	93	80-120						
Lead	1.00	0.90	90	80-120						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **20**
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37623	05/30/2006	H&A

Method: 8082, Polychlorinated Biphenyls (PCBs) by GC

QUALITY CONTROL REPORT

QC Batch No: 060106 Sample Spiked: 060106 QC Prepared: 06/01/2006 QC Analyzed: 06/02/2006 Units: ug/L

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Aroclor-1260 (PCB-1260)	0.0	2.50	2.18	87	2.50	2.10	84	3.5	75-125	<20

QC Batch No: 060106 Sample Spiked: 060106 QC Prepared: 06/01/2006 QC Analyzed: 06/02/2006 Units: ug/L

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Aroclor-1260 (PCB-1260)	2.50	2.15	86	75-125						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 21
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37623	05/30/2006	H&A

Method: 8260B, Volatile Organic Compounds by GC/MS (SW846)

QUALITY CONTROL REPORT

QC Batch No: 060206 Sample Spiked: 060206 QC Prepared: 06/02/2006 QC Analyzed: 06/02/2006 Units: ug/L

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Benzene	0.0	50.00	53.00	106	50.00	54.00	108	1.9	75-125	<20
Chlorobenzene	0.0	50.00	52.00	104	50.00	47.50	95	9.0	75-125	<20
1,1-Dichloroethene	0.0	50.00	53.00	106	50.00	52.50	105	<1	75-125	<20
Methyl-tert-butyl ether (MTBE)	0.0	50.00	57.00	114	50.00	51.00	102	11.1	75-125	<20
Toluene (Methyl benzene)	0.0	50.00	51.00	102	50.00	48.00	96	6.1	75-125	<20
Trichloroethene	0.0	50.00	52.50	105	50.00	54.00	108	2.8	75-125	<20

QC Batch No: 060206 Sample Spiked: 060206 QC Prepared: 06/02/2006 QC Analyzed: 06/02/2006 Units: ug/L

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Benzene	50.00	53.50	107	75-125						
Chlorobenzene	50.00	49.00	98	75-125						
1,1-Dichloroethene	50.00	51.50	103	75-125						
Methyl-tert-butyl ether (MTBE)	50.00	50.50	101	75-125						
Toluene (Methyl benzene)	50.00	49.00	98	75-125						
Trichloroethene	50.00	51.00	102	75-125						
LCS										
Chloroform (Trichloromethane)	50.00	57.00	114	75-125						
Ethylbenzene	50.00	48.00	96	75-125						
1,1,1-Trichloroethane	50.00	57.50	115	75-125						
o-Xylene	50.00	49.50	99	75-125						
m,p-Xylenes	100.00	96.00	96	75-125						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 22
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37623	05/30/2006	H&A

Method: (6010BSCAN), As, Sb,and Pb Metals by ICP

QUALITY CONTROL REPORT

QC Batch No: 050206 Sample Spiked: 37623.01 QC Prepared: 05/02/2006 QC Analyzed: 06/02/2006 Units: mg/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Antimony	0.130	1.00	1.00	87	1.00	1.01	88	1.1	80-120	<15
Arsenic	0.116	1.00	1.03	91	1.00	1.03	91	<1	80-120	<15
Lead	2.100	1.00	3.24	114	1.00	3.27	117	2.6	80-120	<15

QC Batch No: 050206 Sample Spiked: 37623.01 QC Prepared: 05/02/2006 QC Analyzed: 06/02/2006 Units: mg/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Antimony	1.00	0.85	85	80-120						
Arsenic	1.00	0.89	89	80-120						
Lead	1.00	0.88	88	80-120						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 23
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37623	05/30/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QUALITY CONTROL REPORT

QC Batch No: 060106 Sample Spiked: 060106 QC Prepared: 06/01/2006 QC Analyzed: 06/01/2006 Units: ug/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Aroclor-1260 (PCB-1260)	0.0	500.00	425.00	85	500.00	430.00	86	1.2	75-125	<20

QC Batch No: 060106 Sample Spiked: 060106 QC Prepared: 06/01/2006 QC Analyzed: 06/01/2006 Units: ug/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Aroclor-1260 (PCB-1260)	500.00	455.00	91	75-125						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **24**
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37623	05/30/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QUALITY CONTROL REPORT

QC Batch No: 060206 Sample Spiked: 060206 QC Prepared: 06/02/2006 QC Analyzed: 06/02/2006 Units: ug/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Benzene	0.0	50.00	61.00	122	50.00	60.50	121	<1	75-125	<20
Chlorobenzene	0.0	50.00	50.50	101	50.00	57.00	114	12.1	75-125	<20
1,1-Dichloroethene	0.0	50.00	60.50	121	50.00	62.00	124	2.4	75-125	<20
Methyl-tert-butyl ether (MTBE)	0.0	50.00	61.50	123	50.00	59.50	119	3.3	75-125	<20
Toluene (Methyl benzene)	0.0	50.00	51.00	102	50.00	57.00	114	11.1	75-125	<20
Trichloroethene	0.0	50.00	60.00	120	50.00	59.00	118	1.7	75-125	<20

QC Batch No: 060206 Sample Spiked: 060206 QC Prepared: 06/02/2006 QC Analyzed: 06/02/2006 Units: ug/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Benzene	50.00	61.00	122	75-125						
Chlorobenzene	50.00	57.00	114	75-125						
1,1-Dichloroethene	50.00	60.00	120	75-125						
Methyl-tert-butyl ether (MTBE)	50.00	60.00	120	75-125						
Toluene (Methyl benzene)	50.00	58.00	116	75-125						
Trichloroethene	50.00	61.00	122	75-125						
LCS										
Chloroform (Trichloromethane)	50.00	62.00	124	75-125						
Ethylbenzene	50.00	57.50	115	75-125						
1,1,1-Trichloroethane	50.00	61.50	123	75-125						
o-Xylene	50.00	55.00	110	75-125						
m,p-Xylenes	100.00	109.00	109	75-125						

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

Data Qualifier:

*:	In the QC section, sample results have been taken directly from the ICP reading. No preparation factor has been applied.
B:	Analyte was present in the Method Blank.
D:	Result is from a diluted analysis.
E:	Result is beyond calibration limits and is estimated.
H:	Analysis was performed over the allowed holding time due to circumstances which were beyond laboratory control.
J:	Analyte was detected . However, the analyte concentration is an estimated value, which is between the Method Detection Limit (MDL) and the Practical Quantitation Limit (PQL).
M:	Matrix spike recovery is outside control limits due to matrix interference. Laboratory Control Sample recovery was acceptable.
S6:	Surrogate recovery is outside control limits due to matrix interference.
S8:	The analysis of the sample required a dilution such that the surrogate concentration was diluted below the method acceptance criteria.
X:	Results represent LCS and LCSD data.

Definition:

%Limi:	Percent acceptable limits.
%REC:	Percent recovery.
Con.L:	Acceptable Control Limits
Conce:	Added concentration to the sample.
LCS:	Laboratory Control Sample
MDL:	Method Detection Limit is a statistically derived number which is specific for each instrument, each method, and each compound. It indicates a distinctively detectable quantity with 99% probability.
MS:	Matrix Spike
MS DU:	Matrix Spike Duplicate

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

ND: Analyte was not detected in the sample at or above MDL.

PQL: Practical Quantitation Limit or ML (Minimum Level as per RWQCB) is the minimum concentration that can be quantified with more than 99% confidence. Taking into account all aspects of the entire analytical instrumentation and practice.

Recov: Recovered concentration in the sample.

RPD: Relative Percent Difference

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Ordered By

Haley & Aldrich
500 South Kraemer Blvd. Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attention: Tom Tatnall

Number of Pages 7
Date Received 05/30/2006
Date Reported 06/13/2006

Job Number	Order Date	Client
37778	06/08/2006	H&A

Project ID: 32486-006
Project Name: Delphi System Corp.
Site: 1201 N. Magnolia Avenue
Anaheim, CA 92801

RECEIVED
JUN 20 2006
Haley & Aldrich, Inc.

Enclosed please find results of analyses of 12 soil samples which were analyzed as specified on the attached chain of custody. If there are any questions, please do not hesitate to call.

Checked By: _____ Approved By:

Cyrus Razmara, Ph.D.
Laboratory Director

American Environmental Testing Laboratory Inc.
2834 North Naomi Street Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840

CHAIN OF CUSTODY RECORD

№ 38058

AETL JOB No. # 37778 ~~37621~~

COMPANY: Haley + Aldrich
PHONE: 714-985-3434
PROJECT MANAGER: Tom Tatnall
FAX:
PROJECT NAME: Delphi
PROJECT #: 32486-006
SITE NAME AND ADDRESS: 1201 N. Magnolia Ave. Anaheim

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	PCB's 8082	Hold		PCB's 8082							TEST INSTRUCTIONS & COMMENTS
																		*-Req 6/8 72hrs TAT
	CS0253_CC	000-01	5-26-06	0845	Concrete	1"X8" Core	—	X										37621.01
2	CS0254-CC	000.01	↓	0855	↓	↓		↓										37621.02
3	CS0255-CC	000.01		0900														37621.03
4	CS0256-CC	000-01		0905														37621.04
5	DP0253B-SS	004-01		0945	Soil	(1) 4oz WM Jar												37621.05
6	GS0050-SS	000-01		0950														37621.06
7		001-01		0952														37621.07
8		002-01		0954				X										37621.08
9		003-01		0956					X									37621.09
10		004-01		0958					X									37621.10
11	GS0051-SS	000-01		1024				X										37621.11
12		001-01		1026														37621.12
13		002-01		1028														37621.13
14		003-01		1030					X		*							~~37621.14~~ 37778.01
15	GS0052-SS	000-01	↓	1042	↓	↓		X										37621.15

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y/ N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☐ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☒ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsko	Signature:	Signature: Thomas
Printed Name: Glenn Androsko	Printed Name:	Printed Name: Thomas Robinson
Date: 5-30-06 Time: 1620	Date: Time:	Date: 05-30-06 Time: 1830
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: Thomas Robinson	Printed Name:	Printed Name: Argish
Date: 05-30-06 Time: 1625	Date: Time:	Date: 5/30-06 Time: 1830

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 North Naomi Street Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840

CHAIN OF CUSTODY RECORD

№ 38059

AETL JOB No. 37778 ~~#37621~~

COMPANY: Haley & Aldrich
PHONE:
PROJECT MANAGER: Tom Tatnall
FAX:
PROJECT NAME: Delphi
PROJECT #: 32486-006
SITE NAME AND ADDRESS:

SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	EPA 8082 PCB's	Hold		PCBs 8082							TEST INSTRUCTIONS & COMMENTS
																	*-Req 6/8 72hr TAT
1 GS0052_SS	001.01	5-26-06	1044	Soil	#(1) 4oz WMJar		X										37621.16
2	-002-01		1046				X										37621.17
3	-003.01		1048					X		*							~~37621.18~~ 37778.02
4	004.01		1050					X		*							~~37621.19~~ 37778.03
5 GS0053_SS	000.01		1140				X										37621.20
6	-000.02		1141				X										37621.21
7	-001-01		1142				X										37621.22
8	-002-01		1144				X										37621.23
9	-003.01		1146					X									37621.24
10	004.01		1148					X									37621.25
11 GS0054_SS	000.01		1152				X										37621.26
12	-000.02		1153				X										37621.27
13	-001.01		1154				X										37621.28
14	-002.01		1156				X										37621.29
15	-003.01		1158					X									37621.30

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y/ N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☐ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☐ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsko	Signature:	Signature: Thomas
Printed Name: Glenn Androsko	Printed Name:	Printed Name: THOMAS ROBINSON
Date: 5-30-06 Time: 1620	Date: Time:	Date: 05-30-06 Time: 1830
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: Thomas Robinson	Printed Name:	Printed Name: Agisht
Date: 05-30-06 Time: 1625	Date: Time:	Date: 5/30-06 Time: 1830

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

CHAIN OF CUSTODY RECORD

№ 40048

AETL JOB No. ~~#37622~~ 37778

COMPANY: Haley & Aldrich
PROJECT MANAGER: Tom Tatnall
COMPANY ADDRESS:
PHONE:
FAX:
PROJECT NAME: Delphi
PROJECT #: 32486-004
SITE NAME AND ADDRESS:
PO #:

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	Hold	EPA 8082 PCB's	6010 Pb, As, Sb	PCB's 8082						TEST INSTRUCTIONS & COMMENTS
																	*-Req 6/8 72hrs TAT
1	GS0054_SS_	004.01	5-26-06	1200	Soil	(1) 4oz WM Jar		X									37622.01
2	GS0043_SS_	005_01		1235					X								37622.02
3		006_01		1236					X								37622.03
4		007_01		1238				*			*						~~37622.04~~ 37778.04
5		008_01		1240				*			*						~~37622.05~~ 37778.05
6	DP0156B_SS_	005_01		1329					X	X							37622.06
7		005_02		1330					X	X							37622.07
8		006_01		1331					X	[illegible]							37622.08
9		007_01		1333				*			*						~~37622.09~~ 37778.06
10		008_01		1335				X									37622.10
11	DP0156C_SS_	005_01		1401					X								37622.11
12		006_01		1403					X								37622.12
13		007_01		1405				X									37622.13
14	GS0055_SS_	000_01		1414					X								37622.14
15	"	001_01		1416					X								37622.15

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y / N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☐ NORMAL ☒ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☒ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsko	Signature:	Signature: Thomas
Printed Name: Glenn Androsko	Printed Name:	Printed Name: THOMAS ROBINSON
Date: 5-30-06 Time: 1620	Date: Time:	Date: 05-30-06 Time: 1830
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: THOMAS ROBINSON	Printed Name:	Printed Name: Agisht
Date: 05-30-06 Time: 1625	Date: Time:	Date: 5/30-06 Time: 1830

DISTRIBUTION: WHITE - Laboratory. CANARY - Laboratory. PINK - Project/Account Manager. YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 North Naomi Street Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840

CHAIN OF CUSTODY RECORD

№ 38380

AETL JOB No. # 37778 ~~37623~~

COMPANY: Haley & Aldrich
PHONE: 714-985-3434
PROJECT MANAGER: Tom Tatnall
FAX:
PROJECT NAME: Delphi
PROJECT #: 32486-006
SITE NAME AND ADDRESS: 1201 N. Magnolia Ave, Anaheim

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	PCB's 8082	Pb, As, Sb 6010	Hold	PCBs 8082						TEST INSTRUCTIONS & COMMENTS
																	*-Req 6/8 72hrs TAT
1	GS0056_SS_	000-01	5-30-06	1220	Soil	1) 4oz WMTSC	–	X	X								37623.01
2		001-01		1222				X	X								37623.02
3		002-01		1224				X	X								37623.03
4		003-01		1226						X							37623.04
5		005-01		1230						X							37623.05
6	GS0057_SS_	000-01		1240				X	X								37623.06
7		001-01		1242				X	X								37623.07
8		002-01		1244				X	X								37623.08
9		003-01		1246						X							37623.09
10		005-01		1250						X							37623.10
11	GS0058_SS_	000-01		1300				X	X								37623.11
12		001-01		1301				X	X								37623.12
13		001-02		1302				X	X								37623.13
14		002-01		1304				X	X								37623.14
15		003-01		1306						~~X~~	*						~~37623.15~~ 37778-07

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y/ N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☐ NORMAL ☒ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☒ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsko	Signature:	Signature: Thomas
Printed Name: Glenn Androsko	Printed Name:	Printed Name: Thomas Robinson
Date: 5-30-06 Time: 1610	Date: Time:	Date: 05-30-06 Time: 1830
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: Thomas Robinson	Printed Name:	Printed Name: [illegible]
Date: 05-30-06 Time: 1625	Date: Time:	Date: 5/30-06 Time: 1830

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 North Naomi Street Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840

CHAIN OF CUSTODY RECORD

№ 38381

AETL JOB No. 37778 ~~#37623~~

COMPANY: Haley & Aldrich
PHONE:
PROJECT MANAGER: Tom Tatnall
FAX:
PROJECT NAME: Delphi
PROJECT #: 32486-002
SITE NAME AND ADDRESS:

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	Hold	PCB's 8082	Pb, As, Sb Cam	VOC's 8260		PCB's 8082					TEST INSTRUCTIONS & COMMENTS
																		* Req 6/8 72hrs TAT
1	GS0058_SS	005_01	5-30-06	1310	Soil	(1)4oz wm Jar		*					*					~~37623.16~~ 37778.08
2	GS0059_SS	000_01		1320					X	X								37623.17
3		001_01		1322					X	X								37623.18
4		002_01		1324					X	X								37623.19
5		003_01		1326				X										37623.20
6		005_01		1330				X										37623.21
7	GS0060_SS	000_01		1400		(1)4oz wm Jar & VOAs			X	X	X							37623.22
8		001_01		1402					X	X	X							37623.23
9		002_01		1404					X	X	X							37623.24
10		003_01		1406				*					*					~~37623.25~~ 37778.09
11		004_01		1410				*					*					~~37623.26~~ 37778.10
12	GS0061_SS	000_01		1436		(1)4oz wm Jar			X	X								37623.27
13		001_01		1438					X	X								37623.28
14		002_01		1440					X	X								37623.29
15		003_01		1442				*					*					~~37623.30~~ 37778.11

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y/ N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☐ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☒ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Rudusher	Signature:	Signature: Thomas
Printed Name:	Printed Name:	Printed Name: Thomas Robinson
Date: 5-30-06 Time: 1620	Date: Time:	Date: 05-30-06 Time: 1830
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: Thomas Robinson	Printed Name:	Printed Name: Agisht
Date: 05-30-06 Time: 1625	Date: Time:	Date: 5/30-06 Time: 1820

DISTRIBUTION: WHITE - Laboratory CANARY - Laboratory PINK - Project/Account Manager YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

CHAIN OF CUSTODY RECORD

№ 44853

AETL JOB No. 37778 ~~# 37623~~

COMPANY: Haley & Aldrich
PROJECT MANAGER: Tom Tatnall
COMPANY ADDRESS:
PHONE:
FAX:
PROJECT NAME: Delphi
PROJECT #: 32486-006
SITE NAME AND ADDRESS:
PO #:

ANALYSIS REQUESTED

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	Hold	8260 VOC's	8082 PCB's	Pb, As, Sb 6010	PCBs 8082					TEST INSTRUCTIONS & COMMENTS
																	* - Rsq 6/8 72hrs TAT
1	GS0061_SS_005.01		5-30-06	1443	Soil			X				*					~~37623.31~~ 37778.12
2	007.01		↓	1445	"			X									37623.32
3	EB-053006.01			1515	H2O				X	X	X						37623.33
4	TB-053006			—					X								37623.34
5																	
6																	
7																	
8																	
9																	
10																	
11																	
12																	
13																	
14																	
15																	

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y/ N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☐ NORMAL ☒ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☒ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Anderson	Signature:	Signature: Thomas
Printed Name:	Printed Name:	Printed Name: THOMAS ROBINSON
Date: 5-30-06 Time: 1620	Date: Time:	Date: 05-30-06 Time: 1830
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: THOMAS ROBINSON	Printed Name:	Printed Name: Aright
Date: 05-30-06 Time: 1625	Date: Time:	Date: 5/30-06 Time: 1830

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

Cyrus Razmara

From: Androsko, Glenn [GAndrosko@haleyaldrich.com]
Sent: Thursday, June 08, 2006 5:23 PM
To: cyrus@aetlab.com
Cc: Tatnall, Thomas; Hoggan, Kelly
Subject: Delphi additional testing

Cyrus,

Please run the below samples for EPA-8082 / PCB's (72 Hr TAT):

	AETL JOB#
GS0043_SS_007_01	37622
GS0043_SS_008_01	
GS0051_SS_003_01	37621
GS0052_SS_003_01	
GS0052_SS_004_01	
GS0058_SS_003_01	37623
GS0058_SS_005_01	
GS0060_SS_003_01	
GS0060_SS_004_01	
GS0061_SS_003_01	
GS0061_SS_005_01	

Please call if you have any questions.

Thanks,

Glenn Androsko
Field Representative

Haley & Aldrich, Inc.
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Direct: (714) 984-2108
Cell: (714) 709-3721
Fax: (714) 984-2158
Reception: (714) 985-3434
www.haleyaldrich.com
gandrosko@haleyaldrich.com

NOTICE

The information contained in this communication is confidential and privileged proprietary information intended only for the individual or entity to whom it is addressed. Any unauthorized use, distribution, copying or disclosure of this communication is prohibited. If you have received this communication in

Cyrus Razmara

From: Androsko, Glenn [GAndrosko@haleyaldrich.com]
Sent: Thursday, June 08, 2006 5:29 PM
To: cyrus@aetlab.com
Subject: One more

DP0156B_SS_007_01 # 37622

Glenn Androsko
Field Representative

Haley & Aldrich, Inc.
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Direct: (714) 984-2108
Cell: (714) 709-3721
Fax: (714) 984-2158
Reception: (714) 985-3434
www.haleyaldrich.com
gandrosko@haleyaldrich.com

NOTICE

The information contained in this communication is confidential and privileged proprietary information intended only for the individual or entity to whom it is addressed. Any unauthorized use, distribution, copying or disclosure of this communication is prohibited. If you have received this communication in error, please contact the sender immediately.

TABLE
Summary of Soil Analytical Results for Polychlorinated Biphenyls (PCBs) by GC
1201 N. Magnolia Avenue Anaheim, CA 92801

Polychlorinated Biphenyls (PCBs) by GC. Method: (8082)

Sample ID	Units	GS0043-SS-007-01	GS0043-SS-008-01	GS0051-SS-003-01	GS0052-SS-003-01	GS0052-SS-004-01	GS0058-SS-003-01	GS0058-SS-005-01	GS0060-SS-003-01	GS0060-SS-004-01	GS0061-SS-003-01	GS0061-SS-005-01	GS0156B-SS-007-01
Sample Date		5/26/2006	5/26/2006	5/26/2006	5/26/2006	5/26/2006	5/30/2006	5/30/2006	5/30/2006	5/30/2006	5/30/2006	5/30/2006	5/26/2006
Laboratory Job Number		37778	37778	37778	37778	37778	37778	37778	37778	37778	37778	37778	37778
Aroclor-1016 (PCB-1016)	ug/Kg	ND<25.0	ND<25.0	ND<1250	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0
Aroclor-1221 (PCB-1221)	ug/Kg	ND<25.0	ND<25.0	ND<1250	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0
Aroclor-1232 (PCB-1232)	ug/Kg	ND<25.0	ND<25.0	ND<1250	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0
Aroclor-1242 (PCB-1242)	ug/Kg	ND<25.0	ND<25.0	ND<1250	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0
Aroclor-1248 (PCB-1248)	ug/Kg	**198**	**731**	**59,600**	**72.0**	**42.4J**	**1,820**	**664**	**88.4**	ND<25.0	**1,010**	**888**	**49.1J**
Aroclor-1254 (PCB-1254)	ug/Kg	ND<25.0	ND<25.0	ND<1250	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0
Aroclor-1260 (PCB-1260)	ug/Kg	ND<25.0	ND<25.0	ND<1250	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0
Aroclor-1262 (PCB-1262)	ug/Kg	ND<25.0	ND<25.0	ND<1250	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0
Aroclor-1268 (PCB-1268)	ug/Kg	ND<25.0	ND<25.0	ND<1250	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0	ND<25.0

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: SO_8082.xls

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 2
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37778	05/30/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 060906

Our Lab I.D.			Method Blank				
Client Sample I.D.							
Date Sampled							
Date Prepared			06/09/2006				
Preparation Method			3550B				
Date Analyzed			06/09/2006				
Matrix			Soil				
Units			ug/Kg				
Dilution Factor			1				
Analytes	**MDL**	**PQL**	**Results**				
Aroclor-1016 (PCB-1016)	25.0	50.0	ND				
Aroclor-1221 (PCB-1221)	25.0	50.0	ND				
Aroclor-1232 (PCB-1232)	25.0	50.0	ND				
Aroclor-1242 (PCB-1242)	25.0	50.0	ND				
Aroclor-1248 (PCB-1248)	25.0	50.0	ND				
Aroclor-1254 (PCB-1254)	25.0	50.0	ND				
Aroclor-1260 (PCB-1260)	25.0	50.0	ND				
Aroclor-1262 (PCB-1262)	25.0	50.0	ND				
Aroclor-1268 (PCB-1268)	25.0	50.0	ND				

Our Lab I.D.							
Surrogates	**%Rec.Limit**		**% Rec.**				
Tetrachloro-m-xylene	30-150		101				

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 3
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37778	05/30/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 060906

Our Lab I.D.			**37778.01**
Client Sample I.D.			GS0051-SS-0 03-01
Date Sampled			05/26/2006
Date Prepared			06/09/2006
Preparation Method			3550B
Date Analyzed			06/12/2006
Matrix			Soil
Units			ug/Kg
Dilution Factor			50
Analytes	**MDL**	**PQL**	**Results**
Aroclor-1016 (PCB-1016)	1250	2500	ND
Aroclor-1221 (PCB-1221)	1250	2500	ND
Aroclor-1232 (PCB-1232)	1250	2500	ND
Aroclor-1242 (PCB-1242)	1250	2500	ND
Aroclor-1248 (PCB-1248)	1250	2500	59,600
Aroclor-1254 (PCB-1254)	1250	2500	ND
Aroclor-1260 (PCB-1260)	1250	2500	ND
Aroclor-1262 (PCB-1262)	1250	2500	ND
Aroclor-1268 (PCB-1268)	1250	2500	ND

Our Lab I.D.			**37778.01**
Surrogates	**%Rec.Limit**		**% Rec.**
Tetrachloro-m-xylene	30-150		0 S8

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **4**
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37778	05/30/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 060906

Our Lab I.D.			37778.02	37778.03	37778.04	37778.05	37778.06
Client Sample I.D.			GS0052-SS-003-01	GS0052-SS-004-01	GS0043-SS-007-01	GS0043-SS-008-01	GS0156B-SS-007-01
Date Sampled			05/26/2006	05/26/2006	05/26/2006	05/26/2006	05/26/2006
Date Prepared			06/09/2006	06/09/2006	06/09/2006	06/09/2006	06/09/2006
Preparation Method			3550B	3550B	3550B	3550B	3550B
Date Analyzed			06/09/2006	06/09/2006	06/09/2006	06/09/2006	06/09/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
Aroclor-1016 (PCB-1016)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1221 (PCB-1221)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1232 (PCB-1232)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1242 (PCB-1242)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1248 (PCB-1248)	25.0	50.0	72.0	42.4J	198	731	49.1J
Aroclor-1254 (PCB-1254)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1260 (PCB-1260)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1262 (PCB-1262)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1268 (PCB-1268)	25.0	50.0	ND	ND	ND	ND	ND

Our Lab I.D.			37778.02	37778.03	37778.04	37778.05	37778.06
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**
Tetrachloro-m-xylene	30-150		127	135	122	134	128

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 5
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37778	05/30/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 060906

Our Lab I.D.			37778.07	37778.08	37778.09	37778.10	37778.11
Client Sample I.D.			GS0058-SS-0 03-01	GS0058-SS-0 05-01	GS0060-SS-0 03-01	GS0060-SS-0 04-01	GS0061-SS-0 03-01
Date Sampled			05/30/2006	05/30/2006	05/30/2006	05/30/2006	05/30/2006
Date Prepared			06/09/2006	06/09/2006	06/09/2006	06/09/2006	06/09/2006
Preparation Method			3550B	3550B	3550B	3550B	3550B
Date Analyzed			06/09/2006	06/09/2006	06/09/2006	06/09/2006	06/09/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
Aroclor-1016 (PCB-1016)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1221 (PCB-1221)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1232 (PCB-1232)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1242 (PCB-1242)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1248 (PCB-1248)	25.0	50.0	1,820	664	88.4	ND	1,010
Aroclor-1254 (PCB-1254)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1260 (PCB-1260)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1262 (PCB-1262)	25.0	50.0	ND	ND	ND	ND	ND
Aroclor-1268 (PCB-1268)	25.0	50.0	ND	ND	ND	ND	ND

Our Lab I.D.			37778.07	37778.08	37778.09	37778.10	37778.11
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**
Tetrachloro-m-xylene	30-150		142	134	123	128	134

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 6
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37778	05/30/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 060906

Our Lab I.D.			**37778.12**
Client Sample I.D.			GS0061-SS-0 05-01
Date Sampled			05/30/2006
Date Prepared			06/09/2006
Preparation Method			3550B
Date Analyzed			06/09/2006
Matrix			Soil
Units			ug/Kg
Dilution Factor			1
Analytes	**MDL**	**PQL**	**Results**
Aroclor-1016 (PCB-1016)	25.0	50.0	ND
Aroclor-1221 (PCB-1221)	25.0	50.0	ND
Aroclor-1232 (PCB-1232)	25.0	50.0	ND
Aroclor-1242 (PCB-1242)	25.0	50.0	ND
Aroclor-1248 (PCB-1248)	25.0	50.0	888
Aroclor-1254 (PCB-1254)	25.0	50.0	ND
Aroclor-1260 (PCB-1260)	25.0	50.0	ND
Aroclor-1262 (PCB-1262)	25.0	50.0	ND
Aroclor-1268 (PCB-1268)	25.0	50.0	ND

Our Lab I.D.			**37778.12**
Surrogates	**%Rec.Limit**		**% Rec.**
Tetrachloro-m-xylene	30-150		130

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 7
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37778	05/30/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QUALITY CONTROL REPORT

QC Batch No: 060906 Sample Spiked: 060906 QC Prepared: 06/09/2006 QC Analyzed: 06/09/2006 Units: ug/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Aroclor-1260 (PCB-1260)	0.0	500.00	520.00	104	500.00	540.00	108	3.8	50-150	<50

QC Batch No: 060906 Sample Spiked: 060906 QC Prepared: 06/09/2006 QC Analyzed: 06/09/2006 Units: ug/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Aroclor-1260 (PCB-1260)	500.00	520.00	104	50-150						

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

Data Qualifier:

*:	In the QC section, sample results have been taken directly from the ICP reading. No preparation factor has been applied.
B:	Analyte was present in the Method Blank.
D:	Result is from a diluted analysis.
E:	Result is beyond calibration limits and is estimated.
H:	Analysis was performed over the allowed holding time due to circumstances which were beyond laboratory control.
J:	Analyte was detected . However, the analyte concentration is an estimated value, which is between the Method Detection Limit (MDL) and the Practical Quantitation Limit (PQL).
M:	Matrix spike recovery is outside control limits due to matrix interference. Laboratory Control Sample recovery was acceptable.
S6:	Surrogate recovery is outside control limits due to matrix interference.
S8:	The analysis of the sample required a dilution such that the surrogate concentration was diluted below the method acceptance criteria.
X:	Results represent LCS and LCSD data.

Definition:

%Limi:	Percent acceptable limits.
%REC:	Percent recovery.
Con.L:	Acceptable Control Limits
Conce:	Added concentration to the sample.
LCS:	Laboratory Control Sample
MDL:	Method Detection Limit is a statistically derived number which is specific for each instrument, each method, and each compound. It indicates a distinctively detectable quantity with 99% probability.
MS:	Matrix Spike
MS DU:	Matrix Spike Duplicate

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

ND: Analyte was not detected in the sample at or above MDL.

PQL: Practical Quantitation Limit or ML (Minimum Level as per RWQCB) is the minimum concentration that can be quantified with more than 99% confidence. Taking into account all aspects of the entire analytical instrumentation and practice.

Recov: Recovered concentration in the sample.

RPD: Relative Percent Difference

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

RECEIVED

JUN 21 2006

Haley & Aldrich, Inc.

Ordered By

Haley & Aldrich 500 South Kraemer Blvd. Suite 370 Brea, CA 92821-6723	Number of Pages 5 Date Received 03/30/2006 Date Reported 06/16/2006

Telephone: (714)985-3434
Attention: Tom Tatnall

Job Number	Order Date	Client
37848	06/14/2006	H&A

Project ID: 32486-006
Project Name: Delphi System Corp.
Site: 1201 N. Magnolia Avenue
Anaheim, CA 92801

Enclosed please find results of analyses of 30 soil samples which were analyzed as specified on the attached chain of custody. If there are any questions, please do not hesitate to call.

Checked By: _____ Approved By:

Cyrus Razmara, Ph.D.
Laboratory Director

Chain of Custody Record

HALEY & ALDRICH

Job # ~~36806~~ 37848

Client	Project Manager	Laboratory Name	Date	Custody Number
Haley & Aldrich	Tom Tatnall	AETL	3/28/2006	HA-000055
Address	**Telephone Number (Area Code)**	**Laboratory Address & Telephone Number**	**Lab Number**	
500 S. Kraemer Blvd., Suite 370	(714) 985-3434	2834 Naomi St., Burbank CA 91504 Tel: (818) 845-8200		Page 2 of 7

City	State	ZipCode	Site Contact	LabContact
Brea	CA	92821	T. Tatnall	Jim Lin

Project Name and Location (State): Delphi, Anaheim CA

Carrier/Waybill:

Contract/Purchase Order/Quote No.:

Special Instructions/ Conditions of Receipt: (X) due on 06/16/06 48 e mail

Sample /. D. No. and Description (Containers for each sample may be combin on one line)	Date	Time	Matrix: Air	Water	Sed.	Soil	Preservatives Containers: Unpres.	H2SO4	HNO3	HCL	NaOH	ZnAc/NaOH	Analysis: PCB (EPA 8082)	Hold	6010B (A)									Special Instructions
DP0186_SS_007_01	3/28/2006	14:04				X	X							X										36806.13
DP0223_SS_003_01	3/28/2006	08:24				X	X							X	(X)					37848.01				~~36806.14~~
DP0222_SS_002_01	3/28/2006	08:07				X	X						X											36806.15
DP0223_SS_002_01	3/28/2006	08:23				X	X						X											36806.16
DP0186_SS_005_01	3/28/2006	14:03				X	X							X										36806.17
DP0223_SS_001_02	3/28/2006	08:22				X	X						X											36806.18
DP0223_SS_000_01	3/28/2006	08:20				X	X						X		(X)					37848.02				~~36806.19~~
DP0222_SS_001_01	3/28/2006	08:06				X	X						X											36806.20
DP0224_SS_000_01	3/28/2006	08:30				X	X						X											36806.21
DP0223_SS_001_01	3/28/2006	08:21				X	X						X		(X)					37848.03				~~36806.22~~
DP0222_SS_000_01	3/28/2006	08:05				X	X						X											36806.23
DP0222_SS_003_01	3/28/2006	08:08				X	X							X										36806.24

Possible Hazard Identification: ☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown

Sample Disposal: ☐ Return To Client ☒ Disposal By Lab ☐ Archive For ____ Months (A fee may be assessed if samples are retained Months longer than 1 month)

Turn Around Time Required: ☐ 24 Hours ☐ 48 Hours ☐ 7 Days ☐ 14 Days ☐ 21 Days ☒ Others 5 days

QC Requirements (Specify):

		Date	Time			Date	Time
1. Relinquished By	[signature]	3-28-06	1700	1. Received By	[signature]	3-28-06	1700
2. Relinquished By				2. Received By			
3. Relinquished By	[signature]	3/28/06	1930	3. Received By	[signature]	03/30/06	19:30

Comments:

Chain of Custody Record

HALEY & ALDRICH

Job # ~~36806~~ 37848

Client	Project Manager	Laboratory Name	Date	Custody Number
Haley & Aldrich	Tom Tatnall	AETL	3/28/2006	HA-000056
Address	**Telephone Number (Area Code)**	**Laboratory Address & Telephone Number**	**Lab Number**	
500 S. Kraemer Blvd., Suite 370	(714) 985-3434	2834 Naomi St., Burbank CA 91504 Tel: (818) 845-8200		Page 3 of 7

City	State	ZipCode	Site Contact	LabContact
Brea	CA	92821	T. Tatnall	Jim Lin

Project Name and Location (State): Delphi, Anaheim CA

Carrier/Waybill:

Contract/Purchase Order/Quote No.:

Analysis (Attach list if more space is needed)

Sample /. D. No. and Description (Containers for each sample may be combin on one line)	Date	Time	Matrix: Air	Water	Sed.	Soil	Preservatives Containers: Unpres.	H2SO4	HNO3	HCL	NaOH	ZnAc/NaOH	PCB (EPA 8082)	Hold		6010B (As)								Special Instructions/ Conditions of Receipt
DP0227_SS_000_01	3/28/2006	08:55				X	X						X											36806.25
DP0225_SS_000_01	3/28/2006	08:35				X	X						X			X			37848.04					~~36806.26~~
DP0224_SS_002_01	3/28/2006	08:32				X	X							X										36806.27
DP0224_SS_001_01	3/28/2006	08:31				X	X						X											36806.28
DP0225_SS_003_01	3/28/2006	08:37				X	X							X										36806.29
DP0226_SS_003_01	3/28/2006	08:48				X	X							X										36806.30
DP0226_SS_000_01	3/28/2006	08:45				X	X						X											36806.31
DP0225_SS_001_01	3/28/2006	08:36				X	X						X			X			37848.05					~~36806.32~~
DP0226_SS_001_01	3/28/2006	08:46				X	X						X											36806.33
DP0225_SS_002_01	3/28/2006	08:37				X	X							X		X			37848.06					~~36806.34~~
DP0226_SS_002_01	3/28/2006	08:47				X	X							X										36806.35
DP0224_SS_003_01	3/28/2006	08:33				X	X							X										36806.36

Possible Hazard Identification: ☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown

Sample Disposal: ☐ Return To Client ☒ Disposal By Lab ☐ Archive For ____ Months (A fee may be assessed if samples are retained Months longer than 1 month)

Turn Around Time Required: ☐ 24 Hours ☐ 48 Hours ☐ 7 Days ☐ 14 Days ☐ 21 Days ☒ Others 5 Days

QC Requirements (Specify):

	Date	Time		Date	Time
1. Relinquished By: Helen Andrich	3-28-06	1700	1. Received By: [signature]	3-28-06	1700
2. Relinquished By:			2. Received By:		
3. Relinquished By: [signature]	3/28/06	1930	3. Received By: [signature]	03/30/06	19:30

Comments:

Chain of Custody Record

HALEY & ALDRICH

Job # ~~36807~~ 37848

Client	Project Manager	Laboratory Name	Date	Custody Number
Haley & Aldrich	Tom Tatnall	AETL	3/28/2006	HA-000059
Address	**Telephone Number (Area Code)**	**Laboratory Address & Telephone Number**	**Lab Number**	
500 S. Kraemer Blvd., Suite 370	(714) 985-3434	2834 Naomi St., Burbank CA 91504 Tel: (818) 845-8200		Page 6 of 7

City	State	ZipCode	Site Contact	LabContact
Brea	CA	92821	T. Tatnall	Jim Lin

Project Name and Location (State): Delphi, Anaheim CA

Carrier/Waybill:

Contract/Purchase Order/Quote No.:

Analysis (Attach list if more space is needed)

Special Instructions/ Conditions of Receipt

Sample /. D. No. and Description (Containers for each sample may be combin on one line)	Date	Time	Air	Water	Sed.	Soil		Unpres.	H2SO4	HNO3	HCL	NaOH	ZnAc/NaOH		Metals (EPA 6010B) PB only	Hold		6010B(A)									Special Instructions/ Conditions of Receipt
DP0235_SS_003_01	3/28/2006	10:44				X		X							X			X				37848.07					~~36807.24~~
DP0233_SS_005_01	3/28/2006	10:26				X		X								X											36807.25
DP0235_SS_001_01	3/28/2006	10:42				X		X								X											36807.26
DP0233_SS_004_01	3/28/2006	10:23				X		X								X											36807.27
DP0233_SS_007_01	3/28/2006	10:27				X		X							X												36807.28
DP0235_SS_004_01	3/28/2006	10:45				X		X								X											36807.29
DP0235_SS_007_01	3/28/2006	10:48				X		X							X			X				37848.08					~~36807.30~~
DP0237_SS_000_01	3/28/2006	11:05				X		X							X												36807.31
DP0233_SS_003_01	3/28/2006	10:22				X		X							X												36807.32
DP0235_SS_000_01	3/28/2006	10:41				X		X							X			X				37848.09					~~36807.33~~
DP0235_SS_005_01	3/28/2006	10:47				X		X								X		X				37848.10					~~36807.34~~
DP0235_SS_002_01	3/28/2006	10:43				X		X								X		X				37848.11					~~36807.35~~

Possible Hazard Identification: ☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown

Sample Disposal: ☐ Return To Client ☒ Disposal By Lab ☐ Archive For ____ Months (A fee may be assessed if samples are retained Months longer than 1 month)

Turn Around Time Required: ☐ 24 Hours ☐ 48 Hours ☐ 7 Days ☐ 14 Days ☐ 21 Days ☒ Others 5Days

QC Requirements (Specify):

	Date	Time		Date	Time
1. Relinquished By: Glen Andrich	3-28-06	1700	1. Received By: [signature]	3-28-06	1700
2. Relinquished By:			2. Received By:		
3. Relinquished By: [signature]	3/28/06	1930	3. Received By: [signature]	03/28/06	19:30

Comments:

HALEY & ALDRICH
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821
(714) 985-3434 Fax (714) 985-3433

CHAIN OF CUSTODY FORM

Job # ~~36814~~ 37848

2305

Project Name/Address: Delphi-Anaheim 201 N. Magnolia, Anaheim, CA

Project /PO Number: 32486-006

Project Manager: T. Tatnall

Phone Number:

Sampled by (signature): Glenn Andrasko

Fax Number:

Lab Name: AETL

Lab Address:

Analysis Required

Sample Identification	Lab ID No.	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	Lead 6010B	total		6010B (As)					Special Instructions
XR0223_SS_0020_01		S	Sleeve	1	3/30/06	844			X							36814.16
0025_01						845			X							36814.17
XR0225_SS_0000_01						850		X								36814.18
0005_01						851		X								36814.19
0010_01						852		X								36814.20
0015_01						853			X							36814.21
0020_01						854			X							36814.22
0025_01						855			X							36814.23
XR0213_SS_0000_01						950		X			⊗		37848.12			~~36814.24~~
0005_01						951		X								36814.25
0010_01						952		X			⊗		37848.13			~~36814.26~~
0015_01						953			X							36814.27
0020_01						954			X		⊗		37848.14			~~36814.28~~
0025_01						955			X							36814.29
XR0207_SS_0000_01						1000		X								36814.30

Relinquished By:	Date/Time:	Received By:	Date/Time:
Glenn Andrasko	3-30-06	[illegible]	3/30/06 1525
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By: [illegible]	Date/Time: 3/30/06 1705	Received in ~~Lab~~ By: [illegible]	Date/Time: 03/30/06 17:05

Turnaround Time (Check)
24 Hours ___ 5 Days ✓
48 Hours ___ 10 Days ___
72 Hours ___ Other ___

Sample Integrity (Check):
On Ice ___ Intact ___

HALEY & ALDRICH

500 S. Kraemer Blvd., Suite 370
Brea, CA 92821
(714) 985-3434 Fax (714) 985-3433

CHAIN OF CUSTODY FORM

2308

Job# ~~36815~~ 37848

Project Name/Address: Delphi-Anaheim 1201 N. Magnolia

Project /PO Number: 32486-006

Project Manager: T. Tatnall

Phone Number:

Sampled by (signature): Glenn Andrasker

Fax Number:

Lab Name: AETL

Lab Address:

Analysis Required

Sample Identification	Lab ID No.	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	6010B lead	Hold		6010B (A)					Special Instructions
XR0197-SS-0000-01		S	Sleeve	1	3/30/06	1047		X								36815.01
0005-01						1048		X								36815.02
0010-01						1049		X								36815.03
0015-01						1050			X							36815.04
0015-02						1051			X							36815.05
0020-01						1052			X							36815.06
0025-01						1053			X							36815.07
XR0198-SS-0000-01						1055		X								36815.08
0005-01						1056		X			X		37848.15			~~36815.09~~
0010-01						1057		X								36815.10
0015-01						1058			X		X		37848.16			~~36815.11~~
0020-01						1059			X							36815.12
0025-01						1100			X		X		37848.17			~~36815.13~~
XR0196-SS-0000-01						1145		X								36815.14
" 0005-01						1146		X								36815.15

Relinquished By:	Date/Time:	Received By:	Date/Time:
Glenn Andrasker	3-30-06	[signature]	3/30/06 1525
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By: [signature]	Date/Time: 3/30/06 1705	Received in Lab By: [signature]	Date/Time: 03/30/06 17:05

Turnaround Time (Check)
24 Hours ___ 5 Days ✓
48 Hours ___ 10 Days ___
72 Hours ___ Other ___

Sample Integrity (Check):
On Ice ___ Intact ___

HALEY & ALDRICH
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821
(714) 985-3434 Fax (714) 985-3433

CHAIN OF CUSTODY FORM

2309

Job # ~~36815~~ 37848

Project Name/Address: Delphi - Anaheim 1201 N. Magnolia

Project /PO Number: 32486-006

Project Manager: T. Tathall

Phone Number:

Sampled by (signature): Glenn Andreasker

Fax Number:

Lab Name: AETL

Lab Address:

Analysis Required

Sample Identification	Lab ID No.	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	6010B lead	Hold		(6010B)(As)					Special Instructions
XR0196_SS_0010_01		S	Sleeve	1	3/30/06	1147		X	X##							36815.16
XR0196_SS_0015_01						1148			X							36815.17
0020_01						1149			X							36815.18
0025_01						1150			X							36815.19
XR0212_SS_0000_01						1155		X								36815.20
0005_01						1156		X								36815.21
0010_01						1157		X			X		37848.18			~~36815.22~~
0015_01						1159			X							36815.23
0020_01						1200			X							36815.24
0025_01						1201			X		X		37848.19			~~36815.25~~
0010_02						1158		X								36815.26
XR0209_SS_0000_01						~~1200~~ 1205		X			X		37848.20			~~36815.27~~
0005_01						1206		X								36815.28
0010_01						1207		X			X		37848.21			~~36815.29~~
0015_01						1208			X							36815.30

Relinquished By:	Date/Time:	Received By:	Date/Time:
Glenn Andreasker	3-30-06	[signature]	3/30/06 1525
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By: [signature]	Date/Time: 3/30/06 17:05	Received in Lab By: [signature]	Date/Time: 03/30/06 17:05

Turnaround Time (Check)
24 Hours ___ 5 Days ✓
48 Hours ___ 10 Days ___
72 Hours ___ Other ___

Sample Integrity (Check):
On Ice ___ Intact ___

HALEY & ALDRICH
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821
(714) 985-3434 Fax (714) 985-3433

CHAIN OF CUSTODY FORM

2310

Job #~~36815~~ 37848

Project Name/Address: Delphi – Anaheim, 1201 N. Magnolia

Project /PO Number: 32486-006

Project Manager: T. Tathall

Phone Number:

Sampled by (signature): Glenn Anderson

Fax Number:

Lab Name: AETL

Lab Address:

Analysis Required

Sample Identification	Lab ID No.	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	Lead 6010B	Hold		6010B (Al)					Special Instructions
XR0209_SS_0020_01		S	sleeve	1	3/30/06	1209			X		X		37848.22			~~36815.31~~
" 0025_01						1210			X							36815.32
XR0210_SS_0000_01						1215		X								36815.33
0005_01						1216		X								36815.34
0010_01						1217		X								36815.35
0015_01						1218			X							36815.36
0020_01						1219			X							36815.37
0025_01						1220			X							36815.38
XR0214_SS_0000_01						1230		X								36815.39
0005_01						1231		X								36815.40
0010_01						1232		X								36815.41
0015_01						1233			X							36815.42
0020_01						1234			X							36815.43
0025_01						1235			X							36815.44
XR0211_SS_0000_01						1245		X								36815.45

Relinquished By:	Date/Time:	Received By:	Date/Time:
Glenn Anderson	3-30-06	[signature]	3/30/06 1525
[signature]	3/30/06 1705	Received in Lab By: [signature]	03/30/06 17:05

Turnaround Time (Check): 24 Hours ___ 5 Days ✓ ; 48 Hours ___ 10 Days ___ ; 72 Hours ___ Other ___

Sample Integrity (Check): On Ice ___ Intact ___

HALEY & ALDRICH
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821
(714) 985-3434 Fax (714) 985-3433

CHAIN OF CUSTODY FORM

Job # ~~36835~~ 37848

2315

Project Name/Address: Delphi Anaheim 1201 N. Magnolia

Project /PO Number: 32486-006

Project Manager: T. Tatnall

Phone Number:

Sampled by (signature): K. Hogg

Fax Number:

Lab Name: AETL

Lab Address:

Sample Identification	Lab ID No.	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	6010B lead	Total	6010B (A)						Special Instructions
XR0183-SS-0005-01		S	sleeve	1	3/31/06	1226		X		X			37848.23			~~36835.01~~
0010-01						1227		X								36835.02
0015-01						1228			X	X			37848.24			~~36835.03~~
0020-01						1229			X							36835.04
0025-01						1230			X	X			37848.25			~~36835.05~~
XR0184-SS-0000-01						1235		X								36835.06
0005-01						1236		X								36835.07
0010-01						1237		X								36835.08
0015-01						1238			X							36835.09
0020-01						1239			X							36835.10
0025-01						1240			X							36835.11
0025-02						1241			X							36835.12
XR0185-SS-0000-01						1245		X								36835.13
0005-01						1246		X								36835.14
0010-01						1247		X								36835.15

Relinquished By: Kelly Hogg	Date/Time: 3/31/06 1500	~~Received By:~~ relinquished by [illegible]	Date/Time: 3/31/06 1700
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By: [illegible]	Date/Time: 3/31/06 1500	Received in Lab By: [illegible]	Date/Time: 03/31/06 17:06

Turnaround Time (Check): 24 Hours ___ 5 Days ✓ 48 Hours ___ 10 Days ___ 72 Hours ___ Other ___

Sample Integrity (Check): On Ice ___ Intact ___

HALLY & ALDRICH
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821
(714) 985-3434 Fax (714) 985-3433

CHAIN OF CUSTODY FORM

2439

Job # ~~36967~~ 37848

Project Name/Address: Delphi – Anaheim

Project /PO Number: 32486-006

Project Manager: T. Tathall

Phone Number:

Sampled by (signature): R. Hogg

Fax Number:

Lab Name: AETL

Lab Address:

Analysis Required

Sample Identification	Lab ID No.	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	8082 PCBs	Hold		(A) 6010B					Special Instructions
DP0257_SS_000_01		S	Sleeve	1	4-11-06	1640		X			X		37848.26			~~36967.01~~
001_01						1641		X			X		37848.27			~~36967.02~~
002_01						1642			X		X		37848.28			~~36967.03~~
003_01						1643			X							36967.04
DP0258_SS_000_01						1650		X								36967.05
001_01						1651		X								36967.06
002_01						1653			X							36967.07
003_01						1654			X							36967.08
09_02						1652		X								36967.09
DP0254_SS_000_01						1655		X								36967.10
001_01						1656		X	~~X~~							36967.11
002_01						1657			X							36967.12
003_01						1658			X							36967.13
DP0255_SS_000_01						1705		X			X		37848.29			~~36967.14~~
001_01						1706		X			X		37848.30			~~36967.15~~

Relinquished By:	Date/Time:	Received By:	Date/Time:	Turnaround Time (Check)
R. Hogg	1615 4-12-06	[illegible]	4/12/06 1615	24 Hours ___ 5 Days ✓
Relinquished By:	Date/Time:	Received By:	Date/Time:	48 Hours ✓ 10 Days ___
				72 Hours ___ Other ___
Relinquished By: [illegible]	Date/Time: 4/12/06 1755	Received in Lab By: [illegible]	Date/Time: 04/12/06 17:55	Sample Integrity (Check): On Ice ___ Intact ___

Cyrus Razmara

From: Hoggan, Kelly [KHoggan@haleyaldrich.com]
Sent: Wednesday, June 14, 2006 2:56 PM
To: Cyrus Razmara
Cc: Androsko, Glenn; Tatnall, Thomas
Subject: Delphi Additional Arsenic Analysis

Cyrus, can you please run the following samples for Arsenic 6010B with 48hr TOT.

AETL #36806
DP0225_SS_000_01 36806.26
DP0225_SS_001_01 .32
DP0225_SS_002_01 .34

~~#36922~~
DP0223_SS_000_01 36806.19
DP0223_SS_001_01 .22
DP0223_SS_003_01 .14

AETL # 36967
DP0255_SS_000_01 36967.14
DP0255_SS_001_01 .15
DP0255_SS_002_01 .16
DP0255_SS_003_01 .17

DP0256_SS_000_01 36967.18
DP0256_SS_001_01 .19

DP0257_SS_000_01 36967.01
DP0257_SS_001_01 .02
DP0257_SS_002_01 .03

#37091
GS0043_SS_000_01 37091.01
GS0043_SS_002_01 .03
* GS0043_SS_003_01
* GS0043_SS_004_01

GS0044_SS_000_01 37091.04
GS0044_SS_001_01 .05
GS0044_SS_002_01 .06
* GS0044_SS_003_01

GS0045_SS_000_01 37091.07
GS0045_SS_001_01 .08

 # 37390
GS0046_SS_000_01 37390.20
GS0046_SS_001_01 .21

GS0047_SS_000_01 37390.23
GS0047_SS_001_01 .24

GS0048_SS_000_01 37390.26
GS0048_SS_001_01 .27

GS0049_SS_000_01 37390.29 A ✓
GS0049_SS_001_01 .30 ✓

#37621
GS0051_SS_000_01 37621.11 A ✓
GS0051_SS_001_01 .12 ✓
GS0051_SS_003_01 .14 ✓

GS0052_SS_003_01 37621.18 A ✓
GS0052_SS_004_01 .19 ✓

GS0054_SS_000_01 37621.26 A ✓
GS0054_SS_001_01 .28 ✓

#37623
GS0058_SS_003_01 37623.15 A ✓

Kelly Hoggan
Geologist
Haley Aldrich, Inc.
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821-6723
Direct Line: 714.984.2104
Cell: 714.709.3072
Office: 714.985.3434
Fax: 714.985.3433
KHoggan@HaleyAldrich.com
www.HaleyAldrich.com

Hi Glenn,
Looks like you are missing one digit in these sample IDs. Please clarify.

XR0183_SS_0005_01	36835.01	B
XR0183_SS_0015_01	.03	A
XR0183_SS_0025_01	.05	A
XR0198_SS_0005_01	36815.09	B
XR0198_SS_0015_01	.11	A
XR0198_SS_0025_01	.13	A
XR0209_SS_0000_01	36815.27	B
XR0209_SS_0010_01	.29	B
XR0209_SS_0020_01	.31	A
XR0212_SS_0010_01	36815.22	B
XR0212_SS_0025_01	.25	A
XR0213_SS_0000_01	36814.24	B
XR0213_SS_0010_01	.26	B
XR0213_SS_0020_01	.28	A
DP0235_SS_000_01 (0000)	~~36813.19~~	~~B~~ 37807.33
DP0235_SS_002_01		.35
DP0235_SS_003_01		.24
DP0235_SS_005_01	~~36813.20~~	~~B~~ .34
DP0235_SS_007_01		.30

Cyrus

-----Original Message-----
From: Androsko, Glenn [mailto:GAndrosko@haleyaldrich.com]
Sent: Tuesday, June 13, 2006 4:13 PM
To: cyrus@aetlab.com
Cc: Tatnall, Thomas; Hoggan, Kelly
Subject: Additional testing required

Cyrus,

Please run the below samples for EPA 6010 As with 48 Hr. TAT.

Sample ID	AETL Job #
XR0183_SS_0005_01	36835
XR0183_SS_0015_01	
XR0183_SS_0025_01	
XR0198_SS_0005_01	36815
XR0198_SS_0015_01	
XR0198_SS_0025_01	
XR0209_SS_000_01	
XR0209_SS_001_01	
XR0209_SS_002_01	
XR0212_SS_001_01	

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	TCLP Limit mg/L	TTLC Limit mg/Kg	Units	DP0223-SS-000-01	DP0223-SS-001-01	DP0223-SS-003-01	DP0225-SS-000-01	DP0225-SS-001-01	DP0225-SS-002-01	DP0235-SS-000-01	DP0235-SS-002-01	DP0235-SS-003-01
Sample Date					3/28/2006	3/28/2006	3/28/2006	3/28/2006	3/28/2006	3/28/2006	3/28/2006	3/28/2006	3/28/2006
Laboratory Job Number					37848	37848	37848	37848	37848	37848	37848	37848	37848
Arsenic	5.0	5.0	500	mg/Kg	**19.1**	**7.15**	**1.60J**	**59.5**	ND<0.5	**1.15J**	**4.60J**	ND<0.5	ND<0.5

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: SO_6010BSCAN.xls

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	DP0235-SS-005-01	DP0235-SS-007-01	DP0255-SS-000-01	DP0255-SS-001-01	DP0257-SS-000-01	DP0257-SS-001-01	DP0257-SS-002-01	XR0183-SS-0005-01	XR0183-SS-0015-01	XR0183-SS-0025-01	XR0198-SS-0005-01	XR0198-SS-0015-01
Sample Date		3/28/2006	3/28/2006	3/31/2006	3/31/2006	3/30/2006	3/30/2006	3/31/2006	3/30/2006	3/30/2006	3/30/2006	3/30/2006	3/30/2006
Laboratory Job Number		37848	37848	37848	37848	37848	37848	37848	37848	37848	37848	37848	37848
Arsenic	5.0	**2.70J**	**1.10J**	**8.85**	**18.7**	**14.5**	**16.9**	**2.75J**	**3.00J**	**2.65J**	**4.65J**	**1.15J**	**1.85J**

File: SO_6010BSCAN.xls

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	XR0198-SS-0025-01	XR0209-SS-0000-01	XR0209-SS-0010-01	XR0209-SS-0020-01	XR0212-SS-0010-01	XR0212-SS-0025-01	XR0213-SS-0000-01	XR0213-SS-0010-01	XR0213-SS-0020-01
Sample Date		3/30/2006	3/30/2006	3/30/2006	3/30/2006	3/30/2006	3/30/2006	3/30/2006	3/30/2006	3/30/2006
Laboratory Job Number		37848	37848	37848	37848	37848	37848	37848	37848	37848
Arsenic	5.0	ND<0.5	**2.20J**	**1.65J**	**1.45J**	**2.40J**	ND<0.5	**1.20J**	**1.05J**	**1.55J**

File: SO_6010BSCAN.xls

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L
Sample Date	
Laboratory Job Number	
Arsenic	5.0

File: SO_6010BSCAN.xls

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page 2
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37848	03/30/2006	H&A

Analytes			Arsenic			
Methods of Analyses			(6010BSCAN)			
Date Prepared			06/16/2006			
Date Analyzed			06/16/2006			
Matrix			Soil			
QC Batch Number			061606			
Units			mg/Kg			
Method Detection Limit			0.5			
Practical Quantitation Limit			5.0			
Dilution Factor			1			
Lab ID	**Sample ID**	**Sampled**	**Results**			
37848.01	DP0223-SS-003-01	03/28/2006	1.60J			
37848.02	DP0223-SS-000-01	03/28/2006	19.1			
37848.03	DP0223-SS-001-01	03/28/2006	7.15			
37848.04	DP0225-SS-000-01	03/28/2006	59.5			
37848.05	DP0225-SS-001-01	03/28/2006	ND			
37848.06	DP0225-SS-002-01	03/28/2006	1.15J			
37848.07	DP0235-SS-003-01	03/28/2006	ND			
37848.08	DP0235-SS-007-01	03/28/2006	1.10J			
37848.09	DP0235-SS-000-01	03/28/2006	4.60J			
37848.10	DP0235-SS-005-01	03/28/2006	2.70J			
37848.11	DP0235-SS-002-01	03/28/2006	ND			
37848.12	XR0213-SS-0000-01	03/30/2006	1.20J			
37848.13	XR0213-SS-0010-01	03/30/2006	1.05J			
37848.14	XR0213-SS-0020-01	03/30/2006	1.55J			
37848.15	XR0198-SS-0005-01	03/30/2006	1.15J			
37848.16	XR0198-SS-0015-01	03/30/2006	1.85J			
37848.17	XR0198-SS-0025-01	03/30/2006	ND			
N/A	Method Blank	03/28/2006	ND			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37848	03/30/2006	H&A

Analytes			Arsenic			
Methods of Analyses			(6010BSCAN)			
Date Prepared			06/16/2006			
Date Analyzed			06/16/2006			
Matrix			Soil			
QC Batch Number			061606-1			
Units			mg/Kg			
Method Detection Limit			0.5			
Practical Quantitation Limit			5.0			
Dilution Factor			1			
Lab ID	**Sample ID**	**Sampled**	**Results**			
37848.18	XR0212-SS-0010-01	03/30/2006	2.40J			
37848.19	XR0212-SS-0025-01	03/30/2006	ND			
37848.20	XR0209-SS-0000-01	03/30/2006	2.20J			
37848.21	XR0209-SS-0010-01	03/30/2006	1.65J			
37848.22	XR0209-SS-0020-01	03/30/2006	1.45J			
37848.23	XR0183-SS-0005-01	03/30/2006	3.00J			
37848.24	XR0183-SS-0015-01	03/30/2006	2.65J			
37848.25	XR0183-SS-0025-01	03/30/2006	4.65J			
37848.26	DP0257-SS-000-01	03/30/2006	14.5			
37848.27	DP0257-SS-001-01	03/30/2006	16.9			
37848.28	DP0257-SS-002-01	03/31/2006	2.75J			
37848.29	DP0255-SS-000-01	03/31/2006	8.85			
37848.30	DP0255-SS-001-01	03/31/2006	18.7			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **4**
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37848	03/30/2006	H&A

Method: (6010BSCAN), Arsenic by ICP

QUALITY CONTROL REPORT

QC Batch No: 061606 Sample Spiked: 37848.03 QC Prepared: 06/16/2006 QC Analyzed: 06/16/2006 Units: mg/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Arsenic	0.143	1.00	0.97	83	1.00	0.97	83	<1	80-120	<15

QC Batch No: 061606 Sample Spiked: 37848.03 QC Prepared: 06/16/2006 QC Analyzed: 06/16/2006 Units: mg/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Arsenic	1.00	0.94	94	80-120						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 5
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37848	03/30/2006	H&A

Method: (6010BSCAN), Arsenic by ICP

QUALITY CONTROL REPORT

QC Batch No: 061606-1 Sample Spiked: 37848.18 QC Prepared: 06/16/2006 QC Analyzed: 06/16/2006 Units: mg/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Arsenic	0.048	1.00	0.91	86	1.00	0.91	86	<1	80-120	<15

QC Batch No: 061606-1 Sample Spiked: 37848.18 QC Prepared: 06/16/2006 QC Analyzed: 06/16/2006 Units: mg/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Arsenic	1.00	0.93	93	80-120						

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

Data Qualifier:

*:	In the QC section, sample results have been taken directly from the ICP reading. No preparation factor has been applied.
B:	Analyte was present in the Method Blank.
D:	Result is from a diluted analysis.
E:	Result is beyond calibration limits and is estimated.
H:	Analysis was performed over the allowed holding time due to circumstances which were beyond laboratory control.
J:	Analyte was detected . However, the analyte concentration is an estimated value, which is between the Method Detection Limit (MDL) and the Practical Quantitation Limit (PQL).
M:	Matrix spike recovery is outside control limits due to matrix interference. Laboratory Control Sample recovery was acceptable.
S6:	Surrogate recovery is outside control limits due to matrix interference.
S8:	The analysis of the sample required a dilution such that the surrogate concentration was diluted below the method acceptance criteria.
X:	Results represent LCS and LCSD data.

Definition:

%Limi:	Percent acceptable limits.
%REC:	Percent recovery.
Con.L:	Acceptable Control Limits
Conce:	Added concentration to the sample.
LCS:	Laboratory Control Sample
MDL:	Method Detection Limit is a statistically derived number which is specific for each instrument, each method, and each compound. It indicates a distinctively detectable quantity with 99% probability.
MS:	Matrix Spike
MS DU:	Matrix Spike Duplicate

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

ND: Analyte was not detected in the sample at or above MDL.

PQL: Practical Quantitation Limit or ML (Minimum Level as per RWQCB) is the minimum concentration that can be quantified with more than 99% confidence. Taking into account all aspects of the entire analytical instrumentation and practice.

Recov: Recovered concentration in the sample.

RPD: Relative Percent Difference

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

RECEIVED
JUN 21 2006
Haley & Aldrich, Inc.

Ordered By

Haley & Aldrich
500 South Kraemer Blvd. Suite H370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attention: Tom Tatnall

Number of Pages 5
Date Received 04/12/2006
Date Reported 06/16/2006

Job Number	Order Date	Client
37849	06/14/2006	H&A

Project ID: 32486-006
Project Name: Delphi System Corp.
Site: 1201 N. Magnolia Avenue
Anaheim, CA 92801

Enclosed please find results of analyses of 27 soil samples which were analyzed as specified on the attached chain of custody. If there are any questions, please do not hesitate to call.

Checked By: _____ Approved By:

Cyrus Razmara, Ph.D.
Laboratory Director

HALEY & ALDRICH
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821
(714) 985-3434 Fax (714) 985-3433

CHAIN OF CUSTODY FORM

Job# ~~36967~~ 37849

2440

Project Name/Address: Delphi Anaheim

Project /PO Number: 32486-006

Project Manager: T. Tathall

Phone Number:

Sampled by (signature):

Fax Number:

Lab Name: AETL

Lab Address: ⊗ 48 hr rush Due on 06/16/06

Analysis Required

Sample Identification	Lab ID No.	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	8082 PCBs	8260B VOCs	Hold	6010B (As)					Special Instructions
DP0255_SS_002_01		S	Sleeve	1	4-11-06	1707				X	⊗	,	37849.01			~~36967.16~~
003_01						1708				X	⊗	,	37849.02			~~36967.17~~
DP0256_SS_000_01						1710		X			⊗	,	37849.03			~~36967.18~~
001_01						1711		X			⊗	,	37849.04			~~36967.19~~
002_01						1712				X						36967.20
003_01						1713				X						36967.21
DP0167_SS_000_01			Sleeve VOA	1 4	4-12-06	755			X							36967.22
001_01			Sleeve VOA	1 4		757			X							36967.23
007_01			VOA	4		800			X							36967.24
015_01			VOA	4		815			X							36967.25
023_01			VOA	4		840			X							36967.26
DP0170_SS_~~002_01~~ 002_01			Sleeve VOA	1 4		850			X							36967.27
007_01			VOA	4		900			X							36967.28
015_01			VOA	4		915			X							36967.29
022_01			VOA	4		930			X							36967.30

Relinquished By:	Date/Time:	Received By:	Date/Time:
	4-12-06 1615		4/12/06 1615
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received in Lab By:	Date/Time:
	4/12/06 1755		04/12/06 17:55

Turnaround Time (Check): 24 Hours ___ 5 Days ✓ KH; 48 Hours ✓; 10 Days ___; 72 Hours ___; Other ___

Sample Integrity (Check): On Ice ___ Intact ___

HALEY & ALDRICH
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821
(714) 985-3434 Fax (714) 985-3433

CHAIN OF CUSTODY FORM

2443

Job #~~37091~~ 37849

Project Name/Address: Delphi, 1201 N. Magnolia, Anaheim CA

Project /PO Number: 32486-006

Project Manager: Tom Tatnall

Phone Number:

Sampled by (signature): Glenn Androsko

Fax Number:

Lab Name: AETL

Lab Address: 2834 N. Naomi, Burbank CA

Analysis Required

Sample Identification	Lab ID No.	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	8082 PCB's	6010 SB	Hold	6010B (A)					Special Instructions
GS0043_SS_	000-01	Soil	8 oz wm jar	1	4-19-06	1420	—	X			X		37849.05			~~37091.01~~
↓	001-01	↓	↓	1	↓	1421		↓								37091.02
↓	002-01	↓	↓	1	↓	1422		↓			X		37849.06			~~37091.03~~
GS0044_SS_	000-01	↓	↓	1	↓	1445		↓			X		37849.07			~~37091.04~~
↓	001-01	↓	↓	1	↓	1446		↓			X		37849.08			~~37091.05~~
↓	002-01	↓	↓	1	↓	1447		↓			X		37849.09			~~37091.06~~
GS0045_SS_	000-01	↓	↓	1	↓	1520		↓			X		37849.10			~~37091.07~~
↓	001-01	↓	↓	1	↓	1521		↓			X		37849.11			~~37091.08~~
↓	002-01	↓	↓	1	↓	1522		↓								37091.09
GS0025D_SS_	000-01	Soil	4 oz wm jar	1	4-20-06	0900		X	X							37091.10
↓	001-01	↓	↓	1	↓	0901		X	X							37091.11
↓	002-01	↓	↓	1	↓	0902				X						37091.12
GS0025E_SS_	000-01	↓	↓	1	↓	0930		X	X							37091.13
↓	001-01	↓	↓	1	↓	0931		X	X							37091.14
↓	002-01	↓	↓	1	↓	0932				X						37091.15

Relinquished By:	Date/Time:	Received By:	Date/Time:
Glenn Androsko	4-20-06 / 1515	[signature]	4/20/06 1515
[signature]	4/20/06 1750	[signature]	04/20/06 17:50

Received in Lab By: [signature]

Turnaround Time (Check)
24 Hours ___ 5 Days ___
48 Hours ___ 10 Days ___
72 Hours X Other ___

Sample Integrity (Check):
On Ice ___ Intact ___

American Environmental Testing Laboratory Inc.
2834 North Naomi Street Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840

CHAIN OF CUSTODY RECORD

№ 38020

AETL JOB No. 37849 ~~37390~~

COMPANY: Haley & Aldrich
PHONE: 714-985-3436
PROJECT MANAGER: Tom Tatnall
FAX:
PROJECT NAME: Delphi
PROJECT #:
SITE NAME AND ADDRESS: 1201 N. Magnolia Ave Anaheim

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	PCB 8082	HOLD	6010B (As)	TEST INSTRUCTIONS & COMMENTS
	GS0043-SS-003-01	37390.16	5/12/06	1253	SS	1		X			
2	GS0043-SS-004-01	37390.17		1255	SS	1		~~X~~	X		
3	GS0044-SS-003-01	37390.18		1302	SS	1		X			
4	GS0044-SS-004-01	37390.19		1303	SS	1		~~X~~	X		
5	GS0046-SS-000-01	~~37390.20~~		1316	SS	1		X		X	37849.12
6	GS0046-SS-001-01	~~37390.21~~		1318	SS	1		X		X	37849.13
7	GS0046-SS-002-01	37390.22		1319	SS	1		~~X~~	X		
8	GS0047-SS-000-01	~~37390.23~~		1350	SS	1		X		X	37849.14
9	GS0047-SS-001-01	~~37390.24~~		1352	SS	1		X		X	37849.15
10	GS0047-SS-002-01	37390.25		1354	SS	1		~~X~~	X		
11	GS0048-SS-000-01	~~37390.26~~		1404	SS	1		X		X	37849.16
12	GS0048-SS-001-01	~~37390.27~~		1406	SS	1		X		X	37849.17
13	GS0048-SS-002-01	37390.28		1407	SS	1		~~X~~	X		
14	GS0049-SS-000-01	~~37390.29~~		1416	SS	1		X		X	37849.18
15	GS0049-SS-001-01	~~37390.30~~	✓	1417	SS	1		X		X	37849.19

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS	15	PROPERLY COOLED (Y) N / NA	
CUSTODY SEALS Y / N (NA)		SAMPLES INTACT (Y) N / NA	
RECEIVED IN GOOD COND (Y) N		SAMPLES ACCEPTED (Y) N	

TURN AROUND TIME

☐ NORMAL ☒ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☒ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsiki	Signature:	Signature:
Printed Name: Glenn Androsiki	Printed Name:	Printed Name: Carlos Gonzalez
Date: 5-12-06 Time: 1545	Date: Time:	Date: 5/12/06 Time: 1730

RECEIVED BY: 1.	RECEIVED BY: 2.	RECEIVED BY LABORATORY: 3.
Signature:	Signature:	Signature:
Printed Name: Carlos Gonzalez	Printed Name:	Printed Name:
Date: 5-12-06 Time: 1545	Date: Time:	Date: 5/12/06 Time: 1730

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 North Naomi Street Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840

CHAIN OF CUSTODY RECORD

№ 38058

AETL JOB No. # 37849 ~~37621~~

COMPANY: Haley + Aldrich
PHONE: 714-985-3434
PROJECT MANAGER: Tom Tatnall
FAX:
PROJECT NAME: Delphi
PROJECT #: 32486-006
SITE NAME AND ADDRESS: 1201 N. Magnolia Ave. Anaheim

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	PCB's 8082	Hold		6010B (As)							TEST INSTRUCTIONS & COMMENTS
	CS0253_CC	000-01	5-26-06	0845	Concrete	1"X8" Core	—	X										37621.01
2	CS0254-CC	000.01	↓	0855	↓	↓		↓										37621.02
3	CS0256-CC	000.01		0900														37621.03
4	CS0256-CC-	000-01		0905														37621.04
5	DP0253B_SS	004-01		0945	Soil	(1) 4oz WM Jar												37621.05
6	GS0050-SS-	000-01		0950														37621.06
7		001-01		0952														37621.07
8		002-01		0954				↓										37621.08
9		003-01		0956					X									37621.09
10		004-01		0958					X									37621.10
11	GS0051_SS-	000-01		1024				X			⊗							~~37621.11~~ 37849.20
12		001.01		1026							⊗							~~37621.12~~ 37849.21
13		002-01		1028				↓										37621.13
14		003-01		1030					X		⊗							~~37621.14~~ 37849.22
15	GS0052.SS-	000.01		1042				X										37621.15

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y/ N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☐ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☒ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsko	Signature:	Signature: Thomas
Printed Name: Glenn Androsko	Printed Name:	Printed Name: THOMAS ROBINSON
Date: 5-30-06 Time: 1620	Date: Time:	Date: 05-30-06 Time: 1830
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: THOMAS ROBINSON	Printed Name:	Printed Name: Argisht
Date: 05-30-06 Time: 1625	Date: Time:	Date: 5/30-06 Time: 1830

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 North Naomi Street Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840

CHAIN OF CUSTODY RECORD

№ 38059

AETL JOB No. 37849 ~~#37621~~

COMPANY: Haley & Aldrich
PHONE:
PROJECT MANAGER: Tom Tatnall
FAX:
PROJECT NAME: Delphi
PROJECT #: 32486-006
SITE NAME AND ADDRESS:

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	EPA 8082 PCB's	Hold		6010B (As)							TEST INSTRUCTIONS & COMMENTS
1	GS0052_SS	001.01	5-26-06	1044	Soil	(1) 4oz WM Jar		X										37621.16
2	↓	002-01	↓	1046	↓	↓		X										37621.17
3		003.01		1048					~~X~~		X							~~37621.18~~ 37849.23
4		004.01		1050					~~X~~		X							~~37621.19~~ 37849.24
5	GS0053_SS	000.01		1140				X										37621.20
6	↓	000.02		1141				X										37621.21
7		001-01		1142				X										37621.22
8		002-01		1144				X										37621.23
9		003.01		1146					X									37621.24
10		004.01		1148					X									37621.25
11	GS0054_SS	000.01		1152				X			X							~~37621.26~~ 37849.25
12	↓	000.02		1153				X										37621.27
13		001.01		1154				X			X							~~37621.28~~ 37849.26
14		002.01		1156				X										37621.29
15		003.01	↓	1158	↓	↓			X									37621.30

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y / N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☐ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☐ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsko	Signature:	Signature: Thomas
Printed Name: Glenn Androsko	Printed Name:	Printed Name: Thomas Robinson
Date: 5-30-06 Time: 1620	Date: Time:	Date: 05-30-06 Time: 1830
RECEIVED BY: 1.	**RECEIVED BY:**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: Thomas Robinson	Printed Name:	Printed Name: Agisket
Date: 05-30-06 Time: 1625	Date: Time:	Date: 5/30-06 Time: 1830

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 North Naomi Street Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840

CHAIN OF CUSTODY RECORD

№ 38380

AETL JOB No. ~~# 37623~~ 37849

COMPANY: Haley & Aldrich
PHONE: 714-985-3434
PROJECT MANAGER: Tom Tatnall
FAX:
PROJECT NAME: Delphi
PROJECT #: 32486-006
SITE NAME AND ADDRESS: 1201 N. Magnolia Ave, Anaheim

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	PCB's 8082	Pb, As, Sb 6010	Hold	6010B (As)						TEST INSTRUCTIONS & COMMENTS
1	GS0056_SS_	000-01	5-30-06	1220	Soil	1) 4oz WMTSC	–	X	X								37623.01
2		001-01		1222				X	X								37623.02
3		002-01		1224				X	X								37623.03
4		003-01		1226						X							37623.04
5		005-01		1230						X							37623.05
6	GS0057_SS_	000-01		1240				X	X								37623.06
7		001-01		1242				X	X								37623.07
8		002-01		1244				X	X								37623.08
9		003-01		1246						X							37623.09
10		005-01		1250						X							37623.10
11	GS0058_SS_	000-01		1300				X	X								37623.11
12		001-01		1301				X	X								37623.12
13		001-02		1302				X	X								37623.13
14		002-01		1304				X	X								37623.14
15		003-01		1306						X	X						~~37623.15~~ 37849.27

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y/ N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☐ NORMAL ☒ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☒ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsko	Signature:	Signature: Thomas
Printed Name: Glenn Androsko	Printed Name:	Printed Name: Thomas Robinson
Date: 5-30-06 Time: 1620	Date: Time:	Date: 05-30-06 Time: 1830
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: Thomas Robinson	Printed Name:	Printed Name:
Date: 05-30-06 Time: 1625	Date: Time:	Date: 5/30-06 Time: 1830

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	TCLP Limit mg/L	TTLC Limit mg/Kg	Units	DP0255-SS-002-01	DP0255-SS-003-01	DP0256-SS-000-01	DP0256-SS-001-01	GS0043-SS-000-01	GS0043-SS-002-01	GS0044-SS-000-01	GS0044-SS-001-01	GS0044-SS-002-01
Sample Date					4/11/2006	4/11/2006	4/11/2006	4/11/2006	4/19/2006	4/19/2006	4/19/2006	4/19/2006	4/19/2006
Laboratory Job Number					37849	37849	37849	37849	37849	37849	37849	37849	37849
Arsenic	5.0	5.0	500	mg/Kg	**22.1**	**1.95J**	**17.3**	**7.10**	**1.20J**	**1.90J**	**18.3**	**1.45J**	**1.50J**

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: SO_6010BSCAN.xls

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	GS0045-SS-000-01	GS0045-SS-001-01	GS0046-SS-000-01	GS0046-SS-001-01	GS0047-SS-000-01	GS0047-SS-001-01	GS0048-SS-000-01	GS0048-SS-001-01	GS0049-SS-000-01	GS0049-SS-001-01	GS0051-SS-000-01	GS0051-SS-001-01
Sample Date		4/19/2006	4/19/2006	5/12/2006	5/12/2006	5/12/2006	5/12/2006	5/12/2006	5/12/2006	5/12/2006	5/12/2006	5/26/2006	5/26/2006
Laboratory Job Number		37849	37849	37849	37849	37849	37849	37849	37849	37849	37849	37849	37849
Arsenic	5.0	**15.8**	**1.50J**	**1.65J**	**1.25J**	**4.50J**	**1.40J**	**1.70J**	**1.90J**	**1.35J**	**1.95J**	**2.70J**	**3.50J**

File: SO_6010BSCAN.xls

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	GS0051-SS-003-01	GS0052-SS-003-01	GS0052-SS-004-01	GS0054-SS-000-01	GS0054-SS-001-01	GS0058-SS-003-01
Sample Date		5/26/2006	5/26/2006	5/26/2006	5/26/2006	5/26/2006	5/30/2006
Laboratory Job Number		37849	37849	37849	37849	37849	37849
Arsenic	5.0	**2.70J**	**2.60J**	**2.25J**	**3.40J**	**1.55J**	**2.45J**

File: SO_6010BSCAN.xls

Cyrus Razmara

From: Hoggan, Kelly [KHoggan@haleyaldrich.com]
Sent: Wednesday, June 14, 2006 2:56 PM
To: Cyrus Razmara
Cc: Androsko, Glenn; Tatnall, Thomas
Subject: Delphi Additional Arsenic Analysis

Cyrus, can you please run the following samples for Arsenic 6010B with 48hr TOT.

AETL #36806
DP0225_SS_000_01 36806.26
DP0225_SS_001_01 .32
DP0225_SS_002_01 .34

#~~36922~~
DP0223_SS_000_01 36806.19
DP0223_SS_001_01 .22
DP0223_SS_003_01 .14

AETL # 36967
DP0255_SS_000_01 36967.14
DP0255_SS_001_01 .15
DP0255_SS_002_01 .16
DP0255_SS_003_01 .17

DP0256_SS_000_01 36967.18
DP0256_SS_001_01 .19

DP0257_SS_000_01 36967.01
DP0257_SS_001_01 .02
DP0257_SS_002_01 .03

#37091
GS0043_SS_000_01 37091.01
GS0043_SS_002_01 .03
* GS0043_SS_003_01
* GS0043_SS_004_01

GS0044_SS_000_01. 37091.04
GS0044_SS_001_01 .05
GS0044_SS_002_01. .06
* GS0044_SS_003_01

GS0045_SS_000_01 37091.07
GS0045_SS_001_01 .08

 # 37390
GS0046_SS_000_01 37390.20
GS0046_SS_001_01 .21

GS0047_SS_000_01 37390.23
GS0047_SS_001_01 .24

GS0048_SS_000_01 37390.26
GS0048_SS_001_01 .27

GS0049_SS_000_01· 37390.29 A ✓
GS0049_SS_001_01· .30 ✓

#37621
GS0051_SS_000_01· 37621.11 A ✓
GS0051_SS_001_01· .12 ✓
GS0051_SS_003_01· .14 ✓

GS0052_SS_003_01· 37621.18 A ✓
GS0052_SS_004_01· .19 ✓

GS0054_SS_000_01· 37621.26 A ✓
GS0054_SS_001_01· .28 ✓

#37623
GS0058_SS_003_01· 37623.15 A ✓

Kelly Hoggan
Geologist
Haley Aldrich, Inc.
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821-6723
Direct Line: 714.984.2104
Cell: 714.709.3072
Office: 714.985.3434
Fax: 714.985.3433
KHoggan@HaleyAldrich.com
www.HaleyAldrich.com

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714) 985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page 2
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37849	04/12/2006	H&A

Analytes			Arsenic			
Methods of Analyses			(6010BSCAN)			
Date Prepared			06/16/2006			
Date Analyzed			06/16/2006			
Matrix			Soil			
QC Batch Number			061606-2			
Units			mg/Kg			
Method Detection Limit			0.5			
Practical Quantitation Limit			5.0			
Dilution Factor			1			
Lab ID	**Sample ID**	**Sampled**	**Results**			
37849.01	DP0255-SS-002-01	04/11/2006	22.1			
37849.02	DP0255-SS-003-01	04/11/2006	1.95J			
37849.03	DP0256-SS-000-01	04/11/2006	17.3			
37849.04	DP0256-SS-001-01	04/11/2006	7.10			
37849.05	GS0043-SS-000-01	04/19/2006	1.20J			
37849.06	GS0043-SS-002-01	04/19/2006	1.90J			
37849.07	GS0044-SS-000-01	04/19/2006	18.3			
37849.08	GS0044-SS-001-01	04/19/2006	1.45J			
37849.09	GS0044-SS-002-01	04/19/2006	1.50J			
37849.10	GS0045-SS-000-01	04/19/2006	15.8			
37849.11	GS0045-SS-001-01	04/19/2006	1.50J			
37849.12	GS0046-SS-000-01	05/12/2006	1.65J			
37849.13	GS0046-SS-001-01	05/12/2006	1.25J			
37849.14	GS0047-SS-000-01	05/12/2006	4.50J			
37849.15	GS0047-SS-001-01	05/12/2006	1.40J			
37849.16	GS0048-SS-000-01	05/12/2006	1.70J			
37849.17	GS0048-SS-001-01	05/12/2006	1.90J			
37849.18	GS0049-SS-000-01	05/12/2006	1.35J			
37849.19	GS0049-SS-001-01	05/12/2006	1.95J			
37849.20	GS0051-SS-000-01	05/26/2006	2.70J			
N/A	Method Blank	04/11/2006	ND			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714) 985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page 3
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37849	04/12/2006	H&A

Analytes			Arsenic			
Methods of Analyses			(6010BSCAN)			
Date Prepared			06/16/2006			
Date Analyzed			06/16/2006			
Matrix			Soil			
QC Batch Number			061606-1			
Units			mg/Kg			
Method Detection Limit			0.5			
Practical Quantitation Limit			5.0			
Dilution Factor			1			
Lab ID	**Sample ID**	**Sampled**	**Results**			
37849.21	GS0051-SS-001-01	05/26/2006	3.50J			
37849.22	GS0051-SS-003-01	05/26/2006	2.70J			
37849.23	GS0052-SS-003-01	05/26/2006	2.60J			
37849.24	GS0052-SS-004-01	05/26/2006	2.25J			
37849.25	GS0054-SS-000-01	05/26/2006	3.40J			
37849.26	GS0054-SS-001-01	05/26/2006	1.55J			
37849.27	GS0058-SS-003-01	05/30/2006	2.45J			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **4**
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37849	04/12/2006	H&A

Method: (6010BSCAN), Arsenic by ICP

QUALITY CONTROL REPORT

QC Batch No: 061606-1 Sample Spiked: 37848.18 QC Prepared: 06/16/2006 QC Analyzed: 06/16/2006 Units: mg/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Arsenic	0.048	1.00	0.91	86	1.00	0.91	86	<1	80-120	<15

QC Batch No: 061606-1 Sample Spiked: 37848.18 QC Prepared: 06/16/2006 QC Analyzed: 06/16/2006 Units: mg/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Arsenic	1.00	0.93	93	80-120						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 5
Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37849	04/12/2006	H&A

Method: (6010BSCAN), Arsenic by ICP

QUALITY CONTROL REPORT

QC Batch No: 061606-2 Sample Spiked: 37849.03 QC Prepared: 06/16/2006 QC Analyzed: 06/16/2006 Units: mg/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Arsenic	0.346	1.00	1.19	84	1.00	1.20	85	1.2	80-120	<15

QC Batch No: 061606-2 Sample Spiked: 37849.03 QC Prepared: 06/16/2006 QC Analyzed: 06/16/2006 Units: mg/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Arsenic	1.00	0.88	88	80-120						

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

Data Qualifier:

*:	In the QC section, sample results have been taken directly from the ICP reading. No preparation factor has been applied.
B:	Analyte was present in the Method Blank.
D:	Result is from a diluted analysis.
E:	Result is beyond calibration limits and is estimated.
H:	Analysis was performed over the allowed holding time due to circumstances which were beyond laboratory control.
J:	Analyte was detected . However, the analyte concentration is an estimated value, which is between the Method Detection Limit (MDL) and the Practical Quantitation Limit (PQL).
M:	Matrix spike recovery is outside control limits due to matrix interference. Laboratory Control Sample recovery was acceptable.
S6:	Surrogate recovery is outside control limits due to matrix interference.
S8:	The analysis of the sample required a dilution such that the surrogate concentration was diluted below the method acceptance criteria.
X:	Results represent LCS and LCSD data.

Definition:

%Limi:	Percent acceptable limits.
%REC:	Percent recovery.
Con.L:	Acceptable Control Limits
Conce:	Added concentration to the sample.
LCS:	Laboratory Control Sample
MDL:	Method Detection Limit is a statistically derived number which is specific for each instrument, each method, and each compound. It indicates a distinctively detectable quantity with 99% probability.
MS:	Matrix Spike
MS DU:	Matrix Spike Duplicate

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

ND: Analyte was not detected in the sample at or above MDL.

PQL: Practical Quantitation Limit or ML (Minimum Level as per RWQCB) is the minimum concentration that can be quantified with more than 99% confidence. Taking into account all aspects of the entire analytical instrumentation and practice.

Recov: Recovered concentration in the sample.

RPD: Relative Percent Difference