American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

RECEIVED
JUN 27 2006
Haley & Aldrich, Inc.

Ordered By

Haley & Aldrich
500 South Kraemer Blvd. Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attention: Tom Tatnall

Number of Pages 2
Date Received 04/12/2006
Date Reported 06/21/2006

Job Number	Order Date	Client
37886	06/19/2006	H&A

Project ID: 32486-006
Project Name: Delphi System Corp.
Site: 1201 N. Magnolia Avenue
Anaheim, CA 92801

Enclosed please find results of analyses of 2 soil samples which were analyzed as specified on the attached chain of custody. If there are any questions, please do not hesitate to call.

Checked By: _____ Approved By: _____

Cyrus Razmara, Ph.D.
Laboratory Director

HALEY & ALDRICH
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821
(714) 985-3434 Fax (714) 985-3433

CHAIN OF CUSTODY FORM

Job# ~~36967~~ # 37886

2440

Project Name/Address: Delphi Anaheim

Project /PO Number: 32486-006

Project Manager: T. Tathall

Phone Number:

Sampled by (signature):

Fax Number:

Analysis Required

Lab Name: AETL

Lab Address: * - Request 6/19 48 hrs TAT

Sample Identification	Lab ID No.	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	8082 PCBs	8260B VOCs	Hold	As 6010B					Special Instructions
DP0255-SS-002-01		S	sleeve	1	4-11-06	1707				X						36967.16
003-01						1708				X						36967.17
DP0256-SS-000-01						1710		X								36967.18
001-01						1711		X								36967.19
002-01						1712				~~X~~	*		37886.01			~~36967.20~~
003-01						1713				~~X~~	*		37886.02			~~36967.21~~
DP0167-SS-000-01			sleeve VOA	1 4	4-12-06	755			X							36967.22
001-01			sleeve VOA	4		757			X							36967.23
007-01			VOA	4		800			X							36967.24
015-01			VOA	4		815			X							36967.25
023-01			VOA	4		840			X							36967.26
DP0170-SS-~~002-01~~ 002-01 *H			sleeve VOA	1 4		850			X							36967.27
007-01			VOA	4		900			X							36967.28
015-01			VOA	4		915			X							36967.29
022-01			VOA	4		930			X							36967.30

Relinquished By:	Date/Time:	Received By:	Date/Time:
	4-12-06 1615		4/12/06 1615
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received in Lab By:	Date/Time:
	4/12/06 1755		04/12/06 17:55

Turnaround Time (Check): 24 Hours ___ 5 Days ✓ KH; 48 Hours ✓; 10 Days ___; 72 Hours ___; Other ___

Sample Integrity (Check): On Ice ___ Intact ___

6/19-06 1030

Cyrus Razmara

From: Hoggan, Kelly [KHoggan@haleyaldrich.com]
Sent: Monday, June 19, 2006 9:42 AM
To: Cyrus Razmara
Subject: Additional Delphi As Analysis

Cyrus,

Can you please run the following two samples for Arsenic 6010B from AETL Job number 36967 with 48hr TOT.

DP0256_SS_002_01
DP0256_SS_003_01

Thank you

Kelly Hoggan
Geologist
Haley Aldrich, Inc.
500 S. Kraemer Blvd., Suite 370
Brea, CA 92821-6723
Direct Line: 714.984.2104
Cell: 714.709.3072
Office: 714.985.3434
Fax: 714.985.3433
KHoggan@HaleyAldrich.com
www.HaleyAldrich.com

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	TCLP Limit mg/L	TTLC Limit mg/Kg	Units	DP0256-SS-002-01	DP0256-SS-003-01
Sample Date					4/11/2006	4/11/2006
Laboratory Job Number					37886	37886
Arsenic	5.0	5.0	500	mg/Kg	**3.00J**	**2.50J**

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: SO_6010BSCAN.xls

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Telephone: (714)985-3434
Attn: Tom Tatnall

Page: 2

Project ID: 32486-006
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
37886	04/12/2006	H&A

Method: (6010BSCAN), Arsenic by ICP

QC Batch No: 061906-2

Our Lab I.D.			Method Blank	**37886.01**	**37886.02**		
Client Sample I.D.				DP0256-SS-002-01	DP0256-SS-003-01		
Date Sampled				04/11/2006	04/11/2006		
Date Prepared			06/19/2006	06/19/2006	06/19/2006		
Preparation Method			3050B	3050B	3050B		
Date Analyzed			06/19/2006	06/19/2006	06/19/2006		
Matrix			Soil	Soil	Soil		
Units			mg/Kg	mg/Kg	mg/Kg		
Dilution Factor			1	1	1		
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**		
Arsenic	0.5	5.0	ND	3.00J	2.50J		

QUALITY CONTROL REPORT

QC Batch No: 061906-2 Sample Spiked: 37864.05 QC Prepared: 06/19/2006 QC Analyzed: 06/19/2006

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Arsenic	0.022	1.00	0.93	91	1.00	0.92	90	1.1	80-120	<15

QC Batch No: 061906-2 Sample Spiked: 37864.05 QC Prepared: 06/19/2006 QC Analyzed: 06/19/2006

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Arsenic	1.00	0.93	93	80-120						

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

Data Qualifier:

*:	In the QC section, sample results have been taken directly from the ICP reading. No preparation factor has been applied.
B:	Analyte was present in the Method Blank.
D:	Result is from a diluted analysis.
E:	Result is beyond calibration limits and is estimated.
H:	Analysis was performed over the allowed holding time due to circumstances which were beyond laboratory control.
J:	Analyte was detected . However, the analyte concentration is an estimated value, which is between the Method Detection Limit (MDL) and the Practical Quantitation Limit (PQL).
M:	Matrix spike recovery is outside control limits due to matrix interference. Laboratory Control Sample recovery was acceptable.
S6:	Surrogate recovery is outside control limits due to matrix interference.
S8:	The analysis of the sample required a dilution such that the surrogate concentration was diluted below the method acceptance criteria.
X:	Results represent LCS and LCSD data.

Definition:

%Limi:	Percent acceptable limits.
%REC:	Percent recovery.
Con.L:	Acceptable Control Limits
Conce:	Added concentration to the sample.
LCS:	Laboratory Control Sample
MDL:	Method Detection Limit is a statistically derived number which is specific for each instrument, each method, and each compound. It indicates a distinctively detectable quantity with 99% probability.
MS:	Matrix Spike
MS DU:	Matrix Spike Duplicate

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

ND: Analyte was not detected in the sample at or above MDL.

PQL: Practical Quantitation Limit or ML (Minimum Level as per RWQCB) is the minimum concentration that can be quantified with more than 99% confidence. Taking into account all aspects of the entire analytical instrumentation and practice.

Recov: Recovered concentration in the sample.

RPD: Relative Percent Difference

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

RECEIVED
JUL 28 2006
Haley & Aldrich, Inc.

Ordered By

Haley & Aldrich
500 South Kraemer Blvd. Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attention: Tom Tatnall

Number of Pages	**5**
Date Received	**07/14/2006**
Date Reported	**07/24/2006**

Job Number	Order Date	Client
38201	07/14/2006	H&A

Project ID: 32486-004B
Project Name: Delphi System Corp.
Site: 1201 N. Magnolia Ave.
Anaheim, CA 92801

Enclosed please find results of analyses of 2 soil samples which were analyzed as specified on the attached chain of custody. If there are any questions, please do not hesitate to call.

Checked By: _____ Approved By:

Cyrus Razmara, Ph.D.
Laboratory Director

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

CHAIN OF CUSTODY RECORD

№ 44397

AETL JOB No. 38201

Page 1 of 1

COMPANY	Haley & Aldrich	PROJECT MANAGER	T. Tatnall
COMPANY ADDRESS	500 S. Kraemer Blvd #370	PHONE / FAX	714-985-3434
PROJECT NAME	Delphi	PROJECT #	Delphi
SITE NAME AND ADDRESS	1201 N. Magnolia Ave, Ana.	PO #	32486-004B

ANALYSIS REQUESTED

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	Gold Arsenic	PCB's 8082	TEST INSTRUCTIONS & COMMENTS
1	MW-11_SS_0005_01		7-14-06	1227	Soil	(1) 8oz Jar		X	X	38201.01
2	MW-11_SS_003_01			1231				X	X	38201.02
3										
4										
5										
6										
7										
8										
9										
10										
11										
12										
13										
14										
15										

Report on separate PO#

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS	2	PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y / N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☒ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☐ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsko	Signature:	Signature: Thomas
Printed Name: Glenn Androsko	Printed Name:	Printed Name: Thomas Robinson
Date: 7-14-06 Time: 1515	Date: Time:	Date: 07-14-06 Time: 1645

RECEIVED BY: 1.	RECEIVED BY: 2.	RECEIVED BY LABORATORY: 3.
Signature: Thomas	Signature:	Signature:
Printed Name: Thomas Robinson	Printed Name:	Printed Name: Greg
Date: 07-14-06 Time: 555	Date: Time:	Date: 07/14/06 Time: 16:45

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Ave. Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	TCLP Limit mg/L	TTLC Limit mg/Kg	Units	MW-11-SS-0005-01	MW-11-SS-003-01
Sample Date					7/14/2006	7/14/2006
Laboratory Job Number					38201	38201
Arsenic	5.0	5.0	500	mg/Kg	**1.25J**	**1.15J**

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: SO_6010BSCAN.xls

TABLE
Summary of Soil Analytical Results for Polychlorinated Biphenyls (PCBs) by GC
1201 N. Magnolia Ave. Anaheim, CA 92801

Polychlorinated Biphenyls (PCBs) by GC. Method: (8082)

Sample ID	Units	MW-11-SS-0005-01	MW-11-SS-003-01
Sample Date		7/14/2006	7/14/2006
Laboratory Job Number		38201	38201
Aroclor-1016 (PCB-1016)	ug/Kg	ND<25.0	ND<25.0
Aroclor-1221 (PCB-1221)	ug/Kg	ND<25.0	ND<25.0
Aroclor-1232 (PCB-1232)	ug/Kg	ND<25.0	ND<25.0
Aroclor-1242 (PCB-1242)	ug/Kg	ND<25.0	ND<25.0
Aroclor-1248 (PCB-1248)	ug/Kg	ND<25.0	ND<25.0
Aroclor-1254 (PCB-1254)	ug/Kg	ND<25.0	ND<25.0
Aroclor-1260 (PCB-1260)	ug/Kg	ND<25.0	ND<25.0
Aroclor-1262 (PCB-1262)	ug/Kg	ND<25.0	ND<25.0
Aroclor-1268 (PCB-1268)	ug/Kg	ND<25.0	ND<25.0

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: SO_8082.xls

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Site

1201 N. Magnolia Ave.
Anaheim, CA 92801

Telephone: (714)985-3434
Attn: Tom Tatnall

Page: 2

Project ID: 32486-004B
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38201	07/14/2006	H&A

Method: (6010BSCAN), Arsenic by ICP

QC Batch No: 071706

Our Lab I.D.			Method Blank	**38201.01**	**38201.02**		
Client Sample I.D.				MW-11-SS-0 005-01	MW-11-SS-0 03-01		
Date Sampled				07/14/2006	07/14/2006		
Date Prepared			07/17/2006	07/17/2006	07/17/2006		
Preparation Method			3050B	3050B	3050B		
Date Analyzed			07/17/2006	07/17/2006	07/17/2006		
Matrix			Soil	Soil	Soil		
Units			mg/Kg	mg/Kg	mg/Kg		
Dilution Factor			1	1	1		
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**		
Arsenic	0.5	5.0	ND	1.25J	1.15J		

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Site

1201 N. Magnolia Ave.
Anaheim, CA 92801

Telephone: (714)985-3434
Attn: Tom Tatnall

Page: 3

Project ID: 32486-004B
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38201	07/14/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 071706

Our Lab I.D.			Method Blank	**38201.01**	**38201.02**		
Client Sample I.D.				MW-11-SS-0 005-01	MW-11-SS-0 03-01		
Date Sampled				07/14/2006	07/14/2006		
Date Prepared			07/17/2006	07/17/2006	07/17/2006		
Preparation Method			3550B	3550B	3550B		
Date Analyzed			07/18/2006	07/18/2006	07/18/2006		
Matrix			Soil	Soil	Soil		
Units			ug/Kg	ug/Kg	ug/Kg		
Dilution Factor			1	1	1		
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**		
Aroclor-1016 (PCB-1016)	25.0	50.0	ND	ND	ND		
Aroclor-1221 (PCB-1221)	25.0	50.0	ND	ND	ND		
Aroclor-1232 (PCB-1232)	25.0	50.0	ND	ND	ND		
Aroclor-1242 (PCB-1242)	25.0	50.0	ND	ND	ND		
Aroclor-1248 (PCB-1248)	25.0	50.0	ND	ND	ND		
Aroclor-1254 (PCB-1254)	25.0	50.0	ND	ND	ND		
Aroclor-1260 (PCB-1260)	25.0	50.0	ND	ND	ND		
Aroclor-1262 (PCB-1262)	25.0	50.0	ND	ND	ND		
Aroclor-1268 (PCB-1268)	25.0	50.0	ND	ND	ND		
Our Lab I.D.			Method Blank	**38201.01**	**38201.02**		
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**		
Tetrachloro-m-xylene	30-150		114	134	130		

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Ave.
Anaheim, CA 92801

Page: **4**
Project ID: 32486-004B
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38201	07/14/2006	H&A

Method: (6010BSCAN), Arsenic by ICP

QUALITY CONTROL REPORT

QC Batch No: 071706 Sample Spiked: 38198.01 QC Prepared: 07/17/2006 QC Analyzed: 07/17/2006 Units: mg/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Arsenic	0.017	1.00	1.05	103	1.00	1.08	106	2.9	80-120	<15

QC Batch No: 071706 Sample Spiked: 38198.01 QC Prepared: 07/17/2006 QC Analyzed: 07/17/2006 Units: mg/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Arsenic	1.00	0.96	96	80-120						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Ave.
Anaheim, CA 92801

Page: 5
Project ID: 32486-004B
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38201	07/14/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QUALITY CONTROL REPORT

QC Batch No: 071706 Sample Spiked: 071706 QC Prepared: 07/17/2006 QC Analyzed: 07/18/2006 Units: ug/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Aroclor-1260 (PCB-1260)	0.0	500.00	600.00	120	500.00	605.00	121	<1	50-150	<50

QC Batch No: 071706 Sample Spiked: 071706 QC Prepared: 07/17/2006 QC Analyzed: 07/18/2006 Units: ug/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Aroclor-1260 (PCB-1260)	500.00	580.00	116	50-150						

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

Data Qualifier:

*:	In the QC section, sample results have been taken directly from the ICP reading. No preparation factor has been applied.
B:	Analyte was present in the Method Blank.
D:	Result is from a diluted analysis.
E:	Result is beyond calibration limits and is estimated.
H:	Analysis was performed over the allowed holding time due to circumstances which were beyond laboratory control.
J:	Analyte was detected . However, the analyte concentration is an estimated value, which is between the Method Detection Limit (MDL) and the Practical Quantitation Limit (PQL).
M:	Matrix spike recovery is outside control limits due to matrix interference. Laboratory Control Sample recovery was acceptable.
S6:	Surrogate recovery is outside control limits due to matrix interference.
S8:	The analysis of the sample required a dilution such that the surrogate concentration was diluted below the method acceptance criteria.
X:	Results represent LCS and LCSD data.

Definition:

%Limi:	Percent acceptable limits.
%REC:	Percent recovery.
Con.L:	Acceptable Control Limits
Conce:	Added concentration to the sample.
LCS:	Laboratory Control Sample
MDL:	Method Detection Limit is a statistically derived number which is specific for each instrument, each method, and each compound. It indicates a distinctively detectable quantity with 99% probability.
MS:	Matrix Spike
MS DU:	Matrix Spike Duplicate

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

ND: Analyte was not detected in the sample at or above MDL.

PQL: Practical Quantitation Limit or ML (Minimum Level as per RWQCB) is the minimum concentration that can be quantified with more than 99% confidence. Taking into account all aspects of the entire analytical instrumentation and practice.

Recov: Recovered concentration in the sample.

RPD: Relative Percent Difference

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Ordered By

Haley & Aldrich
500 South Kraemer Blvd. Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attention: Tom Tatnall

Number of Pages	**19**
Date Received	**07/14/2006**
Date Reported	**07/26/2006**

Job Number	Order Date	Client
38202	07/14/2006	H&A

Project ID: 32486-004
Project Name: Delphi System Corp.
Site: 1201 N. Magnolia Avenue
Anaheim, CA 92801

Enclosed please find results of analyses of 1 water and 16 soil samples which were analyzed as specified on the attached chain of custody. If there are any questions, please do not hesitate to call.

Checked By: _____ Approved By:

Cyrus Razmara, Ph.D.
Laboratory Director

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

CHAIN OF CUSTODY RECORD

№ 44398

AETL JOB No. 38202

Page 1 of 2

COMPANY: Haley + Aldrich
PROJECT MANAGER: T. Tattnall
COMPANY ADDRESS: 500 S. Kraemer Blvd #370, Brea CA
PHONE: 714-985-3434
FAX: 714-985-3433
PROJECT NAME: Delphi Anaheim - MW10 MW11
PROJECT #: ~~3259~~ 32486-004
SITE NAME AND ADDRESS: 1201 N. Magnolia Ave, Anaheim
PO #:

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	8260D VOCS	Hold SS Sleeve								TEST INSTRUCTIONS & COMMENTS
1	MW-10_SS_005_01		7-14-06	0830	S	4 VOA		X	X								38202.01
2	MW-10-SS_010_01			0837	S	4VOA, 1 ~~Brass~~ SS sleeve		X	X								38202.02
3	MW-10-SS_015_01			0842	S	4VOA 1 SS		X	X								38202.03
4	MW-10-SS_020_01			0847	S	4VOA		X	X								38202.04
5	MW-10_SS_025_01			0857	S	4VOA, 1 S.S.		X	X								38202.05
6	MW-10_SS_029_01			0910	S	4VOA 1 SS.		X	X								38202.06
7	MW-10_SS_035_01			0915	S	4VOA, 1 SS		X	X								38202.07
8	MW-10_SS_040_01			930	S	4VOA, 1 SS		X	X								38202.08
9	TB_071406			0700	W	2 VOA		X	X								38202.09
10	MW-11_SS_005_01			1245	S	4VOA, 1 SS		X	X								38202.10
11	MW-11_SS_010_01			1247	S	1 SS		X									38202.11
12	MW-11_SS_015_01			1250	S	—		X									38202.12
13	MW-11_SS_020_01			1255	S	1 SS		X									38202.13
14	MW-11_SS_025_01			1258	S			X									38202.14
15	MW-11-SS_030_01			1303	S			X									38202.15

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED (Y) N / NA	
CUSTODY SEALS Y / N (NA)		SAMPLES INTACT (Y) N / NA	
RECEIVED IN GOOD COND. (Y) N		SAMPLES ACCEPTED (Y) N	

TURN AROUND TIME

☒ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☐ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Androsko	Signature:	Signature: Thomas
Printed Name: Glenn Androsko	Printed Name:	Printed Name: Thomas Robinson
Date: 7-14-06 Time: 1515	Date: Time:	Date: 07-14-06 Time: 1645
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: Thomas Robinson	Printed Name:	Printed Name:
Date: 07-14-06 Time: 1575	Date: Time:	Date: 07/14/06 Time: 16:45

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

CHAIN OF CUSTODY RECORD

№ 44396

AETL JOB No. 38202

COMPANY: Haley & Aldrich
PROJECT MANAGER: T. Tatnall
COMPANY ADDRESS: 500 S. Kraemer Blvd #370, Brea CA
PHONE:
FAX:
PROJECT NAME: Delphi
PROJECT #:
SITE NAME AND ADDRESS: 1201 N. Magnolia Ave Anaheim
PO #:

ANALYSIS REQUESTED

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	8260B VOC's	Hold SS Sleeve	TEST INSTRUCTIONS & COMMENTS
1	MW-11_SS	035_01	7-14-06	1310	Soil	4 VOA, 1 SS Sleeve	–	X	X	38202.16
2	MW-11_SS	040_01	"	1322	"	"		X	X	38202.17
3										
4										
5										
6										
7										
8										
9										
10										
11										
12										
13										
14										
15										

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED (Y) N / NA	
CUSTODY SEALS Y / N (NA)		SAMPLES INTACT (Y) N / NA	
RECEIVED IN GOOD COND (Y) N		SAMPLES ACCEPTED (Y) N	

TURN AROUND TIME

☒ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☐ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: Glenn Andrasko	Signature:	Signature: Thomas
Printed Name: Glenn Andrasko	Printed Name:	Printed Name: THOMAS ROBINSON
Date: 7-14-06 Time: 1515	Date: Time:	Date: 07-14-06 Time: 1645
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: Thomas	Signature:	Signature:
Printed Name: THOMAS ROBINSON	Printed Name:	Printed Name: Greg
Date: 07-14-06 Time: 1515	Date: Time:	Date: 07/14/06 Time: 16:45

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

TABLE

Summary of Aqueous Analytical Results for Volatile Organic Compounds by GC/MS (SW846)

1201 N. Magnolia Avenue Anaheim, CA 92801

Volatile Organic Compounds by GC/MS (SW846). Method: 8260B

Sample ID	Units	TB-071406
Sample Date		7/14/2006
Laboratory Job Number		38202
Acetone	ug/L	ND<10
Benzene	ug/L	ND<0.5
Bromobenzene (Phenyl bromide)	ug/L	ND<0.5
Bromochloromethane	ug/L	ND<0.5
Bromodichloromethane	ug/L	ND<0.5
Bromoform (Tribromomethane)	ug/L	ND<2.5
Bromomethane (Methyl bromide)	ug/L	ND<1.5
2-Butanone (MEK)	ug/L	ND<5.0
n-Butylbenzene	ug/L	ND<0.5
sec-Butylbenzene	ug/L	ND<0.5
tert-Butylbenzene	ug/L	ND<0.5
Carbon Disulfide	ug/L	ND<0.5
Carbon tetrachloride	ug/L	ND<0.5
Chlorobenzene	ug/L	ND<0.5
Chloroethane	ug/L	ND<1.5
2-Chloroethyl vinyl ether	ug/L	ND<2.5
Chloroform (Trichloromethane)	ug/L	ND<0.5
Chloromethane (Methyl chloride)	ug/L	ND<1.5
2-Chlorotoluene	ug/L	ND<0.5
4-Chlorotoluene	ug/L	ND<0.5
1,2-Dibromo-3-chloropropane (DBCP)	ug/L	ND<2.5
Dibromochloromethane	ug/L	ND<0.5
1,2-Dibromoethane (EDB)	ug/L	ND<0.5
Dibromomethane	ug/L	ND<0.5
1,2-Dichlorobenzene	ug/L	ND<0.5
1,3-Dichlorobenzene	ug/L	ND<0.5
1,4-Dichlorobenzene	ug/L	ND<0.5
Dichlorodifluoromethane	ug/L	ND<1.5
1,1-Dichloroethane	ug/L	ND<0.5
1,2-Dichloroethane (EDC)	ug/L	ND<0.5
1,1-Dichloroethene	ug/L	ND<0.5

TABLE

Summary of Aqueous Analytical Results for Volatile Organic Compounds by GC/MS (SW846)

1201 N. Magnolia Avenue Anaheim, CA 92801

Volatile Organic Compounds by GC/MS (SW846). Method: 8260B

Sample ID	Units	TB-071406
Sample Date		7/14/2006
Laboratory Job Number		38202
cis-1,2-Dichloroethene	ug/L	ND<0.5
trans-1,2-Dichloroethene	ug/L	ND<0.5
1,2-Dichloropropane	ug/L	ND<0.5
1,3-Dichloropropane	ug/L	ND<0.5
2,2-Dichloropropane	ug/L	ND<0.5
1,1-Dichloropropene	ug/L	ND<0.5
cis-1,3-Dichloropropene	ug/L	ND<0.5
trans-1,3-Dichloropropene	ug/L	ND<0.5
Ethylbenzene	ug/L	ND<0.5
Hexachlorobutadiene	ug/L	ND<1.5
2-Hexanone	ug/L	ND<2.5
Isopropylbenzene	ug/L	ND<0.5
p-Isopropyltoluene	ug/L	ND<0.5
4-Methyl-2-pentanone (MIBK)	ug/L	ND<2.5
Methyl-tert-butyl ether (MTBE)	ug/L	ND<0.5
Methylene chloride (DCM)	ug/L	ND<2.0
Naphthalene	ug/L	ND<0.5
n-Propylbenzene	ug/L	ND<0.5
Styrene	ug/L	ND<0.5
1,1,1,2-Tetrachloroethane	ug/L	ND<0.5
1,1,2,2-Tetrachloroethane	ug/L	ND<0.5
Tetrachloroethene	ug/L	ND<0.5
Toluene (Methyl benzene)	ug/L	ND<0.5
1,2,3-Trichlorobenzene	ug/L	ND<0.5
1,2,4-Trichlorobenzene	ug/L	ND<0.5
1,1,1-Trichloroethane	ug/L	ND<0.5
1,1,2-Trichloroethane	ug/L	ND<0.5
Trichloroethene	ug/L	ND<0.5
Trichlorofluoromethane	ug/L	ND<0.5
1,2,3-Trichloropropane	ug/L	ND<0.5
1,2,4-Trimethylbenzene	ug/L	ND<0.5

TABLE
Summary of Aqueous Analytical Results for Volatile Organic Compounds by GC/MS (SW846)
1201 N. Magnolia Avenue Anaheim, CA 92801

Volatile Organic Compounds by GC/MS (SW846). Method: 8260B

Sample ID	Units	TB-071406
Sample Date		7/14/2006
Laboratory Job Number		38202
1,3,5-Trimethylbenzene	ug/L	ND<0.5
Vinyl Acetate	ug/L	ND<0.5
Vinyl chloride (Chloroethene)	ug/L	ND<0.5
o-Xylene	ug/L	ND<0.5
m,p-Xylenes	ug/L	ND<1.0

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: AQ_8260B.xls

TABLE

Summary of Soil Analytical Results for Volatile Organic Compounds by GC/MS (SW846)

1201 N. Magnolia Avenue Anaheim, CA 92801

Volatile Organic Compounds by GC/MS (SW846). Method: (8260B)

Sample ID	Units	MW-10-SS-005-01	MW-10-SS-010-01	MW-10-SS-015-01	MW-10-SS-020-01	MW-10-SS-025-01	MW-10-SS-029-01	MW-10-SS-035-01	MW-10-SS-040-01	MW-11-SS-005-01	MW-11-SS-010-01	MW-11-SS-015-01	MW-11-SS-020-01
Sample Date		7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006
Laboratory Job Number		38202	38202	38202	38202	38202	38202	38202	38202	38202	38202	38202	38202
Acetone	ug/Kg	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25
Benzene	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0
Bromobenzene (Phenyl bromide)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Bromochloromethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Bromodichloromethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Bromoform (Tribromomethane)	ug/Kg	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25
Bromomethane (Methyl bromide)	ug/Kg	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15
2-Butanone (MEK)	ug/Kg	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25
n-Butylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
sec-Butylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
tert-Butylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Carbon Disulfide	ug/Kg	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25
Carbon tetrachloride	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Chlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Chloroethane	ug/Kg	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15
2-Chloroethyl vinyl ether	ug/Kg	ND<50	ND<50	ND<50	ND<50	ND<50	ND<50	ND<50	ND<50	ND<50	ND<50	ND<50	ND<50
Chloroform (Trichloromethane)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Chloromethane (Methyl chloride)	ug/Kg	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15
2-Chlorotoluene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
4-Chlorotoluene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dibromo-3-chloropropane (DBCP)	ug/Kg	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25
Dibromochloromethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dibromoethane (EDB)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Dibromomethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,3-Dichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,4-Dichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Dichlorodifluoromethane	ug/Kg	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15
1,1-Dichloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dichloroethane (EDC)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1-Dichloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	**23.6**	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0

TABLE
Summary of Soil Analytical Results for Volatile Organic Compounds by GC/MS (SW846)
1201 N. Magnolia Avenue Anaheim, CA 92801

Volatile Organic Compounds by GC/MS (SW846). Method: (8260B)

Sample ID	Units	MW-10-SS-005-01	MW-10-SS-010-01	MW-10-SS-015-01	MW-10-SS-020-01	MW-10-SS-025-01	MW-10-SS-029-01	MW-10-SS-035-01	MW-10-SS-040-01	MW-11-SS-005-01	MW-11-SS-010-01	MW-11-SS-015-01	MW-11-SS-020-01
Sample Date		7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006
Laboratory Job Number		38202	38202	38202	38202	38202	38202	38202	38202	38202	38202	38202	38202
cis-1,2-Dichloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
trans-1,2-Dichloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dichloropropane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,3-Dichloropropane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
2,2-Dichloropropane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1-Dichloropropene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
cis-1,3-Dichloropropene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
trans-1,3-Dichloropropene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Ethylbenzene	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0
Hexachlorobutadiene	ug/Kg	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15
2-Hexanone	ug/Kg	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25
Isopropylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
p-Isopropyltoluene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
4-Methyl-2-pentanone (MIBK)	ug/Kg	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25
Methyl-tert-butyl ether (MTBE)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Methylene chloride (DCM)	ug/Kg	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25
Naphthalene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
n-Propylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Styrene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1,1,2-Tetrachloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1,2,2-Tetrachloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Tetrachloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Toluene (Methyl benzene)	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0
1,2,3-Trichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2,4-Trichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1,1-Trichloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1,2-Trichloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Trichloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Trichlorofluoromethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2,3-Trichloropropane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2,4-Trimethylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0

TABLE
Summary of Soil Analytical Results for Volatile Organic Compounds by GC/MS (SW846)
1201 N. Magnolia Avenue Anaheim, CA 92801

Volatile Organic Compounds by GC/MS (SW846). Method: (8260B)

Sample ID	Units	MW-10-SS-005-01	MW-10-SS-010-01	MW-10-SS-015-01	MW-10-SS-020-01	MW-10-SS-025-01	MW-10-SS-029-01	MW-10-SS-035-01	MW-10-SS-040-01	MW-11-SS-005-01	MW-11-SS-010-01	MW-11-SS-015-01	MW-11-SS-020-01
Sample Date		7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006	7/14/2006
Laboratory Job Number		38202	38202	38202	38202	38202	38202	38202	38202	38202	38202	38202	38202
1,3,5-Trimethylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Vinyl Acetate	ug/Kg	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25	ND<25
Vinyl chloride (Chloroethene)	ug/Kg	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15	ND<15
o-Xylene	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0
m,p-Xylenes	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0	ND<2.0

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: SO_8260B.xls

TABLE

Summary of Soil Analytical Results for Volatile Organic Compounds by GC/MS (SW846)

1201 N. Magnolia Avenue Anaheim, CA 92801

Volatile Organic Compounds by GC/MS (SW846). Method: (

Sample ID	Units	MW-11-SS-025-01	MW-11-SS-030-01	MW-11-SS-035-01	MW-11-SS-040-01
Sample Date		7/14/2006	7/14/2006	7/14/2006	7/14/2006
Laboratory Job Number		38202	38202	38202	38202
Acetone	ug/Kg	ND<25	ND<25	ND<25	ND<25
Benzene	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0
Bromobenzene (Phenyl bromide)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Bromochloromethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Bromodichloromethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Bromoform (Tribromomethane)	ug/Kg	ND<25	ND<25	ND<25	ND<25
Bromomethane (Methyl bromide)	ug/Kg	ND<15	ND<15	ND<15	ND<15
2-Butanone (MEK)	ug/Kg	ND<25	ND<25	ND<25	ND<25
n-Butylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
sec-Butylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
tert-Butylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Carbon Disulfide	ug/Kg	ND<25	ND<25	ND<25	ND<25
Carbon tetrachloride	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Chlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Chloroethane	ug/Kg	ND<15	ND<15	ND<15	ND<15
2-Chloroethyl vinyl ether	ug/Kg	ND<50	ND<50	ND<50	ND<50
Chloroform (Trichloromethane)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Chloromethane (Methyl chloride)	ug/Kg	ND<15	ND<15	ND<15	ND<15
2-Chlorotoluene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
4-Chlorotoluene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dibromo-3-chloropropane (DBCP)	ug/Kg	ND<25	ND<25	ND<25	ND<25
Dibromochloromethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dibromoethane (EDB)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Dibromomethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,3-Dichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,4-Dichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Dichlorodifluoromethane	ug/Kg	ND<15	ND<15	ND<15	ND<15
1,1-Dichloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dichloroethane (EDC)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1-Dichloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0

TABLE

Summary of Soil Analytical Results for Volatile Organic Compounds by GC/MS (SW846)

1201 N. Magnolia Avenue Anaheim, CA 92801

Volatile Organic Compounds by GC/MS (SW846). Method: (

Sample ID	Units	MW-11-SS-025-01	MW-11-SS-030-01	MW-11-SS-035-01	MW-11-SS-040-01
Sample Date		7/14/2006	7/14/2006	7/14/2006	7/14/2006
Laboratory Job Number		38202	38202	38202	38202
cis-1,2-Dichloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
trans-1,2-Dichloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2-Dichloropropane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,3-Dichloropropane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
2,2-Dichloropropane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1-Dichloropropene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
cis-1,3-Dichloropropene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
trans-1,3-Dichloropropene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Ethylbenzene	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0
Hexachlorobutadiene	ug/Kg	ND<15	ND<15	ND<15	ND<15
2-Hexanone	ug/Kg	ND<25	ND<25	ND<25	ND<25
Isopropylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
p-Isopropyltoluene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
4-Methyl-2-pentanone (MIBK)	ug/Kg	ND<25	ND<25	ND<25	ND<25
Methyl-tert-butyl ether (MTBE)	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Methylene chloride (DCM)	ug/Kg	ND<25	ND<25	ND<25	ND<25
Naphthalene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
n-Propylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Styrene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1,1,2-Tetrachloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1,2,2-Tetrachloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Tetrachloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Toluene (Methyl benzene)	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0
1,2,3-Trichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2,4-Trichlorobenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1,1-Trichloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,1,2-Trichloroethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Trichloroethene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Trichlorofluoromethane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2,3-Trichloropropane	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
1,2,4-Trimethylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0

TABLE
Summary of Soil Analytical Results for Volatile Organic Compounds by GC/MS (SW846)
1201 N. Magnolia Avenue Anaheim, CA 92801

Volatile Organic Compounds by GC/MS (SW846). Method: (

Sample ID	Units	MW-11-SS-025-01	MW-11-SS-030-01	MW-11-SS-035-01	MW-11-SS-040-01
Sample Date		7/14/2006	7/14/2006	7/14/2006	7/14/2006
Laboratory Job Number		38202	38202	38202	38202
1,3,5-Trimethylbenzene	ug/Kg	ND<5.0	ND<5.0	ND<5.0	ND<5.0
Vinyl Acetate	ug/Kg	ND<25	ND<25	ND<25	ND<25
Vinyl chloride (Chloroethene)	ug/Kg	ND<15	ND<15	ND<15	ND<15
o-Xylene	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0
m,p-Xylenes	ug/Kg	ND<2.0	ND<2.0	ND<2.0	ND<2.0

File: SO_8260B.xls

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 2
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072406

Our Lab I.D.			Method Blank	38202.01	38202.02	38202.03	38202.04
Client Sample I.D.				MW-10-SS-005-01	MW-10-SS-010-01	MW-10-SS-015-01	MW-10-SS-020-01
Date Sampled				07/14/2006	07/14/2006	07/14/2006	07/14/2006
Date Prepared			07/24/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Preparation Method			5035A	5035A	5035A	5035A	5035A
Date Analyzed			07/24/2006	07/24/2006	07/24/2006	07/24/2006	07/24/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
Acetone	25	50	ND	ND	ND	ND	ND
Benzene	2.0	10.0	ND	ND	ND	ND	ND
Bromobenzene (Phenyl bromide)	5.0	10.0	ND	ND	ND	ND	ND
Bromochloromethane	5.0	10.0	ND	ND	ND	ND	ND
Bromodichloromethane	5.0	10.0	ND	ND	ND	ND	ND
Bromoform (Tribromomethane)	25	50	ND	ND	ND	ND	ND
Bromomethane (Methyl bromide)	15	30	ND	ND	ND	ND	ND
2-Butanone (MEK)	25	50	ND	ND	ND	ND	ND
n-Butylbenzene	5.0	10.0	ND	ND	ND	ND	ND
sec-Butylbenzene	5.0	10.0	ND	ND	ND	ND	ND
tert-Butylbenzene	5.0	10.0	ND	ND	ND	ND	ND
Carbon Disulfide	25	50	ND	ND	ND	ND	ND
Carbon tetrachloride	5.0	10.0	ND	ND	ND	ND	ND
Chlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
Chloroethane	15	30	ND	ND	ND	ND	ND
2-Chloroethyl vinyl ether	50	50	ND	ND	ND	ND	ND
Chloroform (Trichloromethane)	5.0	10.0	ND	ND	ND	ND	ND
Chloromethane (Methyl chloride)	15	30	ND	ND	ND	ND	ND
2-Chlorotoluene	5.0	10.0	ND	ND	ND	ND	ND
4-Chlorotoluene	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dibromo-3-chloropropane (DBCP)	25	50	ND	ND	ND	ND	ND
Dibromochloromethane	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dibromoethane (EDB)	5.0	10.0	ND	ND	ND	ND	ND
Dibromomethane	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,3-Dichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,4-Dichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Page: 3
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072406

Our Lab I.D.			Method Blank	**38202.01**	**38202.02**	**38202.03**	**38202.04**
Client Sample I.D.				MW-10-SS-005-01	MW-10-SS-010-01	MW-10-SS-015-01	MW-10-SS-020-01
Date Sampled				07/14/2006	07/14/2006	07/14/2006	07/14/2006
Date Prepared			07/24/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Preparation Method			5035A	5035A	5035A	5035A	5035A
Date Analyzed			07/24/2006	07/24/2006	07/24/2006	07/24/2006	07/24/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
Dichlorodifluoromethane	15	30	ND	ND	ND	ND	ND
1,1-Dichloroethane	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dichloroethane (EDC)	5.0	10.0	ND	ND	ND	ND	ND
1,1-Dichloroethene	5.0	10.0	ND	ND	ND	ND	ND
cis-1,2-Dichloroethene	5.0	10.0	ND	ND	ND	ND	ND
trans-1,2-Dichloroethene	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dichloropropane	5.0	10.0	ND	ND	ND	ND	ND
1,3-Dichloropropane	5.0	10.0	ND	ND	ND	ND	ND
2,2-Dichloropropane	5.0	10.0	ND	ND	ND	ND	ND
1,1-Dichloropropene	5.0	10.0	ND	ND	ND	ND	ND
cis-1,3-Dichloropropene	5.0	10.0	ND	ND	ND	ND	ND
trans-1,3-Dichloropropene	5.0	10.0	ND	ND	ND	ND	ND
Ethylbenzene	2.0	10.0	ND	ND	ND	ND	ND
Hexachlorobutadiene	15	30	ND	ND	ND	ND	ND
2-Hexanone	25	50	ND	ND	ND	ND	ND
Isopropylbenzene	5.0	10.0	ND	ND	ND	ND	ND
p-Isopropyltoluene	5.0	10.0	ND	ND	ND	ND	ND
4-Methyl-2-pentanone (MIBK)	25	50	ND	ND	ND	ND	ND
Methyl-tert-butyl ether (MTBE)	5.0	10.0	ND	ND	ND	ND	ND
Methylene chloride (DCM)	25	50	ND	ND	ND	ND	ND
Naphthalene	5.0	10.0	ND	ND	ND	ND	ND
n-Propylbenzene	5.0	10.0	ND	ND	ND	ND	ND
Styrene	5.0	10.0	ND	ND	ND	ND	ND
1,1,1,2-Tetrachloroethane	5.0	10.0	ND	ND	ND	ND	ND
1,1,2,2-Tetrachloroethane	5.0	10.0	ND	ND	ND	ND	ND
Tetrachloroethene	5.0	10.0	ND	ND	ND	ND	ND
Toluene (Methyl benzene)	2.0	10.0	ND	ND	ND	ND	ND
1,2,3-Trichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,2,4-Trichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,1,1-Trichloroethane	5.0	10.0	ND	ND	ND	ND	ND
1,1,2-Trichloroethane	5.0	10.0	ND	ND	ND	ND	ND
Trichloroethene	5.0	10.0	ND	ND	ND	ND	ND
Trichlorofluoromethane	5.0	10.0	ND	ND	ND	ND	ND
1,2,3-Trichloropropane	5.0	10.0	ND	ND	ND	ND	ND
1,2,4-Trimethylbenzene	5.0	10.0	ND	ND	ND	ND	ND

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072406

Our Lab I.D.			Method Blank	**38202.01**	**38202.02**	**38202.03**	**38202.04**
Client Sample I.D.				MW-10-SS-005-01	MW-10-SS-010-01	MW-10-SS-015-01	MW-10-SS-020-01
Date Sampled				07/14/2006	07/14/2006	07/14/2006	07/14/2006
Date Prepared			07/24/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Preparation Method			5035A	5035A	5035A	5035A	5035A
Date Analyzed			07/24/2006	07/24/2006	07/24/2006	07/24/2006	07/24/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
1,3,5-Trimethylbenzene	5.0	10.0	ND	ND	ND	ND	ND
Vinyl Acetate	25	50	ND	ND	ND	ND	ND
Vinyl chloride (Chloroethene)	15	30	ND	ND	ND	ND	ND
o-Xylene	2.0	10.0	ND	ND	ND	ND	ND
m,p-Xylenes	2.0	20.0	ND	ND	ND	ND	ND

Our Lab I.D.				**38202.01**	**38202.02**	**38202.03**	**38202.04**
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**
Bromofluorobenzene	75-125		112	107	110	104	111
Dibromofluoromethane	75-125		103	114	114	112	110
Toluene-d8	75-125		112	99	94	92	96

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 5
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072406

Our Lab I.D.			**38202.05**	**38202.06**	**38202.07**		
Client Sample I.D.			MW-10-SS-025-01	MW-10-SS-029-01	MW-10-SS-035-01		
Date Sampled			07/14/2006	07/14/2006	07/14/2006		
Date Prepared			07/14/2006	07/14/2006	07/14/2006		
Preparation Method			5035A	5035A	5035A		
Date Analyzed			07/24/2006	07/24/2006	07/24/2006		
Matrix			Soil	Soil	Soil		
Units			ug/Kg	ug/Kg	ug/Kg		
Dilution Factor			1	1	1		
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**		
Acetone	25	50	ND	ND	ND		
Benzene	2.0	10.0	ND	ND	ND		
Bromobenzene (Phenyl bromide)	5.0	10.0	ND	ND	ND		
Bromochloromethane	5.0	10.0	ND	ND	ND		
Bromodichloromethane	5.0	10.0	ND	ND	ND		
Bromoform (Tribromomethane)	25	50	ND	ND	ND		
Bromomethane (Methyl bromide)	15	30	ND	ND	ND		
2-Butanone (MEK)	25	50	ND	ND	ND		
n-Butylbenzene	5.0	10.0	ND	ND	ND		
sec-Butylbenzene	5.0	10.0	ND	ND	ND		
tert-Butylbenzene	5.0	10.0	ND	ND	ND		
Carbon Disulfide	25	50	ND	ND	ND		
Carbon tetrachloride	5.0	10.0	ND	ND	ND		
Chlorobenzene	5.0	10.0	ND	ND	ND		
Chloroethane	15	30	ND	ND	ND		
2-Chloroethyl vinyl ether	50	50	ND	ND	ND		
Chloroform (Trichloromethane)	5.0	10.0	ND	ND	ND		
Chloromethane (Methyl chloride)	15	30	ND	ND	ND		
2-Chlorotoluene	5.0	10.0	ND	ND	ND		
4-Chlorotoluene	5.0	10.0	ND	ND	ND		
1,2-Dibromo-3-chloropropane (DBCP)	25	50	ND	ND	ND		
Dibromochloromethane	5.0	10.0	ND	ND	ND		
1,2-Dibromoethane (EDB)	5.0	10.0	ND	ND	ND		
Dibromomethane	5.0	10.0	ND	ND	ND		
1,2-Dichlorobenzene	5.0	10.0	ND	ND	ND		
1,3-Dichlorobenzene	5.0	10.0	ND	ND	ND		
1,4-Dichlorobenzene	5.0	10.0	ND	ND	ND		

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Page: 6
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072406

Our Lab I.D.			**38202.05**	**38202.06**	**38202.07**		
Client Sample I.D.			MW-10-SS-0 25-01	MW-10-SS-0 29-01	MW-10-SS-0 35-01		
Date Sampled			07/14/2006	07/14/2006	07/14/2006		
Date Prepared			07/14/2006	07/14/2006	07/14/2006		
Preparation Method			5035A	5035A	5035A		
Date Analyzed			07/24/2006	07/24/2006	07/24/2006		
Matrix			Soil	Soil	Soil		
Units			ug/Kg	ug/Kg	ug/Kg		
Dilution Factor			1	1	1		
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**		
Dichlorodifluoromethane	15	30	ND	ND	ND		
1,1-Dichloroethane	5.0	10.0	ND	ND	ND		
1,2-Dichloroethane (EDC)	5.0	10.0	ND	ND	ND		
1,1-Dichloroethene	5.0	10.0	ND	ND	23.6		
cis-1,2-Dichloroethene	5.0	10.0	ND	ND	ND		
trans-1,2-Dichloroethene	5.0	10.0	ND	ND	ND		
1,2-Dichloropropane	5.0	10.0	ND	ND	ND		
1,3-Dichloropropane	5.0	10.0	ND	ND	ND		
2,2-Dichloropropane	5.0	10.0	ND	ND	ND		
1,1-Dichloropropene	5.0	10.0	ND	ND	ND		
cis-1,3-Dichloropropene	5.0	10.0	ND	ND	ND		
trans-1,3-Dichloropropene	5.0	10.0	ND	ND	ND		
Ethylbenzene	2.0	10.0	ND	ND	ND		
Hexachlorobutadiene	15	30	ND	ND	ND		
2-Hexanone	25	50	ND	ND	ND		
Isopropylbenzene	5.0	10.0	ND	ND	ND		
p-Isopropyltoluene	5.0	10.0	ND	ND	ND		
4-Methyl-2-pentanone (MIBK)	25	50	ND	ND	ND		
Methyl-tert-butyl ether (MTBE)	5.0	10.0	ND	ND	ND		
Methylene chloride (DCM)	25	50	ND	ND	ND		
Naphthalene	5.0	10.0	ND	ND	ND		
n-Propylbenzene	5.0	10.0	ND	ND	ND		
Styrene	5.0	10.0	ND	ND	ND		
1,1,1,2-Tetrachloroethane	5.0	10.0	ND	ND	ND		
1,1,2,2-Tetrachloroethane	5.0	10.0	ND	ND	ND		
Tetrachloroethene	5.0	10.0	ND	ND	ND		
Toluene (Methyl benzene)	2.0	10.0	ND	ND	ND		
1,2,3-Trichlorobenzene	5.0	10.0	ND	ND	ND		
1,2,4-Trichlorobenzene	5.0	10.0	ND	ND	ND		
1,1,1-Trichloroethane	5.0	10.0	ND	ND	ND		
1,1,2-Trichloroethane	5.0	10.0	ND	ND	ND		
Trichloroethene	5.0	10.0	ND	ND	ND		
Trichlorofluoromethane	5.0	10.0	ND	ND	ND		
1,2,3-Trichloropropane	5.0	10.0	ND	ND	ND		
1,2,4-Trimethylbenzene	5.0	10.0	ND	ND	ND		

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Page: 7
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072406

Our Lab I.D.			**38202.05**	**38202.06**	**38202.07**		
Client Sample I.D.			MW-10-SS-025-01	MW-10-SS-029-01	MW-10-SS-035-01		
Date Sampled			07/14/2006	07/14/2006	07/14/2006		
Date Prepared			07/14/2006	07/14/2006	07/14/2006		
Preparation Method			5035A	5035A	5035A		
Date Analyzed			07/24/2006	07/24/2006	07/24/2006		
Matrix			Soil	Soil	Soil		
Units			ug/Kg	ug/Kg	ug/Kg		
Dilution Factor			1	1	1		
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**		
1,3,5-Trimethylbenzene	5.0	10.0	ND	ND	ND		
Vinyl Acetate	25	50	ND	ND	ND		
Vinyl chloride (Chloroethene)	15	30	ND	ND	ND		
o-Xylene	2.0	10.0	ND	ND	ND		
m,p-Xylenes	2.0	20.0	ND	ND	ND		

Our Lab I.D.			**38202.05**	**38202.06**	**38202.07**		
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**		
Bromofluorobenzene	75-125		105	106	106		
Dibromofluoromethane	75-125		111	110	107		
Toluene-d8	75-125		91	94	95		

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 8
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072506

Our Lab I.D.			Method Blank	**38202.08**	**38202.10**	**38202.11**	**38202.12**
Client Sample I.D.				MW-10-SS-0 40-01	MW-11-SS-0 05-01	MW-11-SS-0 10-01	MW-11-SS-0 15-01
Date Sampled				07/14/2006	07/14/2006	07/14/2006	07/14/2006
Date Prepared			07/25/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Preparation Method			5035A	5035A	5035A	5035A	5035A
Date Analyzed			07/25/2006	07/25/2006	07/25/2006	07/25/2006	07/25/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
Acetone	25	50	ND	ND	ND	ND	ND
Benzene	2.0	10.0	ND	ND	ND	ND	ND
Bromobenzene (Phenyl bromide)	5.0	10.0	ND	ND	ND	ND	ND
Bromochloromethane	5.0	10.0	ND	ND	ND	ND	ND
Bromodichloromethane	5.0	10.0	ND	ND	ND	ND	ND
Bromoform (Tribromomethane)	25	50	ND	ND	ND	ND	ND
Bromomethane (Methyl bromide)	15	30	ND	ND	ND	ND	ND
2-Butanone (MEK)	25	50	ND	ND	ND	ND	ND
n-Butylbenzene	5.0	10.0	ND	ND	ND	ND	ND
sec-Butylbenzene	5.0	10.0	ND	ND	ND	ND	ND
tert-Butylbenzene	5.0	10.0	ND	ND	ND	ND	ND
Carbon Disulfide	25	50	ND	ND	ND	ND	ND
Carbon tetrachloride	5.0	10.0	ND	ND	ND	ND	ND
Chlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
Chloroethane	15	30	ND	ND	ND	ND	ND
2-Chloroethyl vinyl ether	50	50	ND	ND	ND	ND	ND
Chloroform (Trichloromethane)	5.0	10.0	ND	ND	ND	ND	ND
Chloromethane (Methyl chloride)	15	30	ND	ND	ND	ND	ND
2-Chlorotoluene	5.0	10.0	ND	ND	ND	ND	ND
4-Chlorotoluene	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dibromo-3-chloropropane (DBCP)	25	50	ND	ND	ND	ND	ND
Dibromochloromethane	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dibromoethane (EDB)	5.0	10.0	ND	ND	ND	ND	ND
Dibromomethane	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,3-Dichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,4-Dichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Page: 9
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072506

Our Lab I.D.			Method Blank	**38202.08**	**38202.10**	**38202.11**	**38202.12**
Client Sample I.D.				MW-10-SS-040-01	MW-11-SS-005-01	MW-11-SS-010-01	MW-11-SS-015-01
Date Sampled				07/14/2006	07/14/2006	07/14/2006	07/14/2006
Date Prepared			07/25/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Preparation Method			5035A	5035A	5035A	5035A	5035A
Date Analyzed			07/25/2006	07/25/2006	07/25/2006	07/25/2006	07/25/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
Dichlorodifluoromethane	15	30	ND	ND	ND	ND	ND
1,1-Dichloroethane	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dichloroethane (EDC)	5.0	10.0	ND	ND	ND	ND	ND
1,1-Dichloroethene	5.0	10.0	ND	ND	ND	ND	ND
cis-1,2-Dichloroethene	5.0	10.0	ND	ND	ND	ND	ND
trans-1,2-Dichloroethene	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dichloropropane	5.0	10.0	ND	ND	ND	ND	ND
1,3-Dichloropropane	5.0	10.0	ND	ND	ND	ND	ND
2,2-Dichloropropane	5.0	10.0	ND	ND	ND	ND	ND
1,1-Dichloropropene	5.0	10.0	ND	ND	ND	ND	ND
cis-1,3-Dichloropropene	5.0	10.0	ND	ND	ND	ND	ND
trans-1,3-Dichloropropene	5.0	10.0	ND	ND	ND	ND	ND
Ethylbenzene	2.0	10.0	ND	ND	ND	ND	ND
Hexachlorobutadiene	15	30	ND	ND	ND	ND	ND
2-Hexanone	25	50	ND	ND	ND	ND	ND
Isopropylbenzene	5.0	10.0	ND	ND	ND	ND	ND
p-Isopropyltoluene	5.0	10.0	ND	ND	ND	ND	ND
4-Methyl-2-pentanone (MIBK)	25	50	ND	ND	ND	ND	ND
Methyl-tert-butyl ether (MTBE)	5.0	10.0	ND	ND	ND	ND	ND
Methylene chloride (DCM)	25	50	ND	ND	ND	ND	ND
Naphthalene	5.0	10.0	ND	ND	ND	ND	ND
n-Propylbenzene	5.0	10.0	ND	ND	ND	ND	ND
Styrene	5.0	10.0	ND	ND	ND	ND	ND
1,1,1,2-Tetrachloroethane	5.0	10.0	ND	ND	ND	ND	ND
1,1,2,2-Tetrachloroethane	5.0	10.0	ND	ND	ND	ND	ND
Tetrachloroethene	5.0	10.0	ND	ND	ND	ND	ND
Toluene (Methyl benzene)	2.0	10.0	ND	ND	ND	ND	ND
1,2,3-Trichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,2,4-Trichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,1,1-Trichloroethane	5.0	10.0	ND	ND	ND	ND	ND
1,1,2-Trichloroethane	5.0	10.0	ND	ND	ND	ND	ND
Trichloroethene	5.0	10.0	ND	ND	ND	ND	ND
Trichlorofluoromethane	5.0	10.0	ND	ND	ND	ND	ND
1,2,3-Trichloropropane	5.0	10.0	ND	ND	ND	ND	ND
1,2,4-Trimethylbenzene	5.0	10.0	ND	ND	ND	ND	ND

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Page: 10
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072506

Our Lab I.D.			Method Blank	**38202.08**	**38202.10**	**38202.11**	**38202.12**
Client Sample I.D.				MW-10-SS-0 40-01	MW-11-SS-0 05-01	MW-11-SS-0 10-01	MW-11-SS-0 15-01
Date Sampled				07/14/2006	07/14/2006	07/14/2006	07/14/2006
Date Prepared			07/25/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Preparation Method			5035A	5035A	5035A	5035A	5035A
Date Analyzed			07/25/2006	07/25/2006	07/25/2006	07/25/2006	07/25/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
1,3,5-Trimethylbenzene	5.0	10.0	ND	ND	ND	ND	ND
Vinyl Acetate	25	50	ND	ND	ND	ND	ND
Vinyl chloride (Chloroethene)	15	30	ND	ND	ND	ND	ND
o-Xylene	2.0	10.0	ND	ND	ND	ND	ND
m,p-Xylenes	2.0	20.0	ND	ND	ND	ND	ND

Our Lab I.D.				**38202.08**	**38202.10**	**38202.11**	**38202.12**
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**
Bromofluorobenzene	75-125		120	108	109	108	107
Dibromofluoromethane	75-125		102	111	113	113	113
Toluene-d8	75-125		108	100	94	93	95

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 11
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072506

Our Lab I.D.			38202.13	38202.14	38202.15	38202.16	38202.17
Client Sample I.D.			MW-11-SS-020-01	MW-11-SS-025-01	MW-11-SS-030-01	MW-11-SS-035-01	MW-11-SS-040-01
Date Sampled			07/14/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Date Prepared			07/14/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Preparation Method			5035A	5035A	5035A	5035A	5035A
Date Analyzed			07/25/2006	07/25/2006	07/25/2006	07/25/2006	07/25/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
Acetone	25	50	ND	ND	ND	ND	ND
Benzene	2.0	10.0	ND	ND	ND	ND	ND
Bromobenzene (Phenyl bromide)	5.0	10.0	ND	ND	ND	ND	ND
Bromochloromethane	5.0	10.0	ND	ND	ND	ND	ND
Bromodichloromethane	5.0	10.0	ND	ND	ND	ND	ND
Bromoform (Tribromomethane)	25	50	ND	ND	ND	ND	ND
Bromomethane (Methyl bromide)	15	30	ND	ND	ND	ND	ND
2-Butanone (MEK)	25	50	ND	ND	ND	ND	ND
n-Butylbenzene	5.0	10.0	ND	ND	ND	ND	ND
sec-Butylbenzene	5.0	10.0	ND	ND	ND	ND	ND
tert-Butylbenzene	5.0	10.0	ND	ND	ND	ND	ND
Carbon Disulfide	25	50	ND	ND	ND	ND	ND
Carbon tetrachloride	5.0	10.0	ND	ND	ND	ND	ND
Chlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
Chloroethane	15	30	ND	ND	ND	ND	ND
2-Chloroethyl vinyl ether	50	50	ND	ND	ND	ND	ND
Chloroform (Trichloromethane)	5.0	10.0	ND	ND	ND	ND	ND
Chloromethane (Methyl chloride)	15	30	ND	ND	ND	ND	ND
2-Chlorotoluene	5.0	10.0	ND	ND	ND	ND	ND
4-Chlorotoluene	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dibromo-3-chloropropane (DBCP)	25	50	ND	ND	ND	ND	ND
Dibromochloromethane	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dibromoethane (EDB)	5.0	10.0	ND	ND	ND	ND	ND
Dibromomethane	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,3-Dichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,4-Dichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072506

Our Lab I.D.			**38202.13**	**38202.14**	**38202.15**	**38202.16**	**38202.17**
Client Sample I.D.			MW-11-SS-0 20-01	MW-11-SS-0 25-01	MW-11-SS-0 30-01	MW-11-SS-0 35-01	MW-11-SS-0 40-01
Date Sampled			07/14/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Date Prepared			07/14/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Preparation Method			5035A	5035A	5035A	5035A	5035A
Date Analyzed			07/25/2006	07/25/2006	07/25/2006	07/25/2006	07/25/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
Dichlorodifluoromethane	15	30	ND	ND	ND	ND	ND
1,1-Dichloroethane	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dichloroethane (EDC)	5.0	10.0	ND	ND	ND	ND	ND
1,1-Dichloroethene	5.0	10.0	ND	ND	ND	ND	ND
cis-1,2-Dichloroethene	5.0	10.0	ND	ND	ND	ND	ND
trans-1,2-Dichloroethene	5.0	10.0	ND	ND	ND	ND	ND
1,2-Dichloropropane	5.0	10.0	ND	ND	ND	ND	ND
1,3-Dichloropropane	5.0	10.0	ND	ND	ND	ND	ND
2,2-Dichloropropane	5.0	10.0	ND	ND	ND	ND	ND
1,1-Dichloropropene	5.0	10.0	ND	ND	ND	ND	ND
cis-1,3-Dichloropropene	5.0	10.0	ND	ND	ND	ND	ND
trans-1,3-Dichloropropene	5.0	10.0	ND	ND	ND	ND	ND
Ethylbenzene	2.0	10.0	ND	ND	ND	ND	ND
Hexachlorobutadiene	15	30	ND	ND	ND	ND	ND
2-Hexanone	25	50	ND	ND	ND	ND	ND
Isopropylbenzene	5.0	10.0	ND	ND	ND	ND	ND
p-Isopropyltoluene	5.0	10.0	ND	ND	ND	ND	ND
4-Methyl-2-pentanone (MIBK)	25	50	ND	ND	ND	ND	ND
Methyl-tert-butyl ether (MTBE)	5.0	10.0	ND	ND	ND	ND	ND
Methylene chloride (DCM)	25	50	ND	ND	ND	ND	ND
Naphthalene	5.0	10.0	ND	ND	ND	ND	ND
n-Propylbenzene	5.0	10.0	ND	ND	ND	ND	ND
Styrene	5.0	10.0	ND	ND	ND	ND	ND
1,1,1,2-Tetrachloroethane	5.0	10.0	ND	ND	ND	ND	ND
1,1,2,2-Tetrachloroethane	5.0	10.0	ND	ND	ND	ND	ND
Tetrachloroethene	5.0	10.0	ND	ND	ND	ND	ND
Toluene (Methyl benzene)	2.0	10.0	ND	ND	ND	ND	ND
1,2,3-Trichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,2,4-Trichlorobenzene	5.0	10.0	ND	ND	ND	ND	ND
1,1,1-Trichloroethane	5.0	10.0	ND	ND	ND	ND	ND
1,1,2-Trichloroethane	5.0	10.0	ND	ND	ND	ND	ND
Trichloroethene	5.0	10.0	ND	ND	ND	ND	ND
Trichlorofluoromethane	5.0	10.0	ND	ND	ND	ND	ND
1,2,3-Trichloropropane	5.0	10.0	ND	ND	ND	ND	ND
1,2,4-Trimethylbenzene	5.0	10.0	ND	ND	ND	ND	ND

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Page: 13
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 072506

Our Lab I.D.			38202.13	38202.14	38202.15	38202.16	38202.17
Client Sample I.D.			MW-11-SS-0 20-01	MW-11-SS-0 25-01	MW-11-SS-0 30-01	MW-11-SS-0 35-01	MW-11-SS-0 40-01
Date Sampled			07/14/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Date Prepared			07/14/2006	07/14/2006	07/14/2006	07/14/2006	07/14/2006
Preparation Method			5035A	5035A	5035A	5035A	5035A
Date Analyzed			07/25/2006	07/25/2006	07/25/2006	07/25/2006	07/25/2006
Matrix			Soil	Soil	Soil	Soil	Soil
Units			ug/Kg	ug/Kg	ug/Kg	ug/Kg	ug/Kg
Dilution Factor			1	1	1	1	1
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**	**Results**	**Results**
1,3,5-Trimethylbenzene	5.0	10.0	ND	ND	ND	ND	ND
Vinyl Acetate	25	50	ND	ND	ND	ND	ND
Vinyl chloride (Chloroethene)	15	30	ND	ND	ND	ND	ND
o-Xylene	2.0	10.0	ND	ND	ND	ND	ND
m,p-Xylenes	2.0	20.0	ND	ND	ND	ND	ND

Our Lab I.D.			38202.13	38202.14	38202.15	38202.16	38202.17
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**	**% Rec.**
Bromofluorobenzene	75-125		106	109	110	112	105
Dibromofluoromethane	75-125		116	114	117	114	11
Toluene-d8	75-125		94	93	92	95	94

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **14**
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38202	07/14/2006	H&A

Method: 8260B, Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 071706

Our Lab I.D.			Method Blank	**38202.09**			
Client Sample I.D.				TB-071406			
Date Sampled				07/14/2006			
Date Prepared			07/17/2006	07/17/2006			
Preparation Method			5030B	5030B			
Date Analyzed			07/17/2006	07/17/2006			
Matrix			Aqueous	Aqueous			
Units			ug/L	ug/L			
Dilution Factor			1	1			
Analytes	**MDL**	**PQL**	**Results**	**Results**			
Acetone	10	10	ND	ND			
Benzene	0.5	1.0	ND	ND			
Bromobenzene (Phenyl bromide)	0.5	1.0	ND	ND			
Bromochloromethane	0.5	1.0	ND	ND			
Bromodichloromethane	0.5	1.0	ND	ND			
Bromoform (Tribromomethane)	2.5	5.0	ND	ND			
Bromomethane (Methyl bromide)	1.5	3.0	ND	ND			
2-Butanone (MEK)	5.0	5.0	ND	ND			
n-Butylbenzene	0.5	1.0	ND	ND			
sec-Butylbenzene	0.5	1.0	ND	ND			
tert-Butylbenzene	0.5	1.0	ND	ND			
Carbon Disulfide	0.5	1.0	ND	ND			
Carbon tetrachloride	0.5	1.0	ND	ND			
Chlorobenzene	0.5	1.0	ND	ND			
Chloroethane	1.5	3.0	ND	ND			
2-Chloroethyl vinyl ether	2.5	5.0	ND	ND			
Chloroform (Trichloromethane)	0.5	1.0	ND	ND			
Chloromethane (Methyl chloride)	1.5	3.0	ND	ND			
2-Chlorotoluene	0.5	1.0	ND	ND			
4-Chlorotoluene	0.5	1.0	ND	ND			
1,2-Dibromo-3-chloropropane (DBCP)	2.5	5.0	ND	ND			
Dibromochloromethane	0.5	1.0	ND	ND			
1,2-Dibromoethane (EDB)	0.5	1.0	ND	ND			
Dibromomethane	0.5	1.0	ND	ND			
1,2-Dichlorobenzene	0.5	1.0	ND	ND			
1,3-Dichlorobenzene	0.5	1.0	ND	ND			
1,4-Dichlorobenzene	0.5	1.0	ND	ND			
Dichlorodifluoromethane	1.5	3.0	ND	ND			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Page: 15
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: 8260B, Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 071706

Our Lab I.D.			Method Blank	**38202.09**			
Client Sample I.D.				TB-071406			
Date Sampled				07/14/2006			
Date Prepared			07/17/2006	07/17/2006			
Preparation Method			5030B	5030B			
Date Analyzed			07/17/2006	07/17/2006			
Matrix			Aqueous	Aqueous			
Units			ug/L	ug/L			
Dilution Factor			1	1			
Analytes	**MDL**	**PQL**	**Results**	**Results**			
1,1-Dichloroethane	0.5	1.0	ND	ND			
1,2-Dichloroethane (EDC)	0.5	1.0	ND	ND			
1,1-Dichloroethene	0.5	1.0	ND	ND			
cis-1,2-Dichloroethene	0.5	1.0	ND	ND			
trans-1,2-Dichloroethene	0.5	1.0	ND	ND			
1,2-Dichloropropane	0.5	1.0	ND	ND			
1,3-Dichloropropane	0.5	1.0	ND	ND			
2,2-Dichloropropane	0.5	1.0	ND	ND			
1,1-Dichloropropene	0.5	1.0	ND	ND			
cis-1,3-Dichloropropene	0.5	1.0	ND	ND			
trans-1,3-Dichloropropene	0.5	1.0	ND	ND			
Ethylbenzene	0.5	1.0	ND	ND			
Hexachlorobutadiene	1.5	3.0	ND	ND			
2-Hexanone	2.5	5.0	ND	ND			
Isopropylbenzene	0.5	1.0	ND	ND			
p-Isopropyltoluene	0.5	1.0	ND	ND			
4-Methyl-2-pentanone (MIBK)	2.5	5.0	ND	ND			
Methyl-tert-butyl ether (MTBE)	0.5	1.0	ND	ND			
Methylene chloride (DCM)	2.0	4.0	ND	ND			
Naphthalene	0.5	1.0	ND	ND			
n-Propylbenzene	0.5	1.0	ND	ND			
Styrene	0.5	1.0	ND	ND			
1,1,1,2-Tetrachloroethane	0.5	1.0	ND	ND			
1,1,2,2-Tetrachloroethane	0.5	1.0	ND	ND			
Tetrachloroethene	0.5	1.0	ND	ND			
Toluene (Methyl benzene)	0.5	1.0	ND	ND			
1,2,3-Trichlorobenzene	0.5	1.0	ND	ND			
1,2,4-Trichlorobenzene	0.5	1.0	ND	ND			
1,1,1-Trichloroethane	0.5	1.0	ND	ND			
1,1,2-Trichloroethane	0.5	1.0	ND	ND			
Trichloroethene	0.5	1.0	ND	ND			
Trichlorofluoromethane	0.5	1.0	ND	ND			
1,2,3-Trichloropropane	0.5	1.0	ND	ND			
1,2,4-Trimethylbenzene	0.5	1.0	ND	ND			
1,3,5-Trimethylbenzene	0.5	1.0	ND	ND			
Vinyl Acetate	0.5	5.0	ND	ND			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Page: 16
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submittd	Client
38202	07/14/2006	H&A

Method: 8260B, Volatile Organic Compounds by GC/MS (SW846)

QC Batch No: 071706

Our Lab I.D.			Method Blank	**38202.09**			
Client Sample I.D.				TB-071406			
Date Sampled				07/14/2006			
Date Prepared			07/17/2006	07/17/2006			
Preparation Method			5030B	5030B			
Date Analyzed			07/17/2006	07/17/2006			
Matrix			Aqueous	Aqueous			
Units			ug/L	ug/L			
Dilution Factor			1	1			
Analytes	**MDL**	**PQL**	**Results**	**Results**			
Vinyl chloride (Chloroethene)	0.5	3.0	ND	ND			
o-Xylene	0.5	1.0	ND	ND			
m,p-Xylenes	1.0	2.0	ND	ND			

Our Lab I.D.				**38202.09**			
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**			
Bromofluorobenzene	75-125		90	101			
Dibromofluoromethane	75-125		93	88			
Toluene-d8	75-125		108	106			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 17
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38202	07/14/2006	H&A

Method: 8260B, Volatile Organic Compounds by GC/MS (SW846)

QUALITY CONTROL REPORT

QC Batch No: 071706 Sample Spiked: 071706 QC Prepared: 07/17/2006 QC Analyzed: 07/17/2006 Units: ug/L

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Benzene	0.00	50.00	47.50	95	50.00	48.00	96	1.0	75-125	<20
Chlorobenzene	0.00	50.00	44.50	89	50.00	46.00	92	3.3	75-125	<20
1,1-Dichloroethene	0.00	50.00	55.50	111	50.00	50.50	101	9.4	75-125	<20
Methyl-tert-butyl ether (MTBE)	0.00	50.00	52.50	105	50.00	48.00	96	9.0	75-125	<20
Toluene (Methyl benzene)	0.00	50.00	43.00	86	50.00	43.50	87	1.2	75-125	<20
Trichloroethene	0.00	50.00	45.00	90	50.00	45.50	91	1.1	75-125	<20

QC Batch No: 071706 Sample Spiked: 071706 QC Prepared: 07/17/2006 QC Analyzed: 07/17/2006 Units: ug/L

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Benzene	50.00	49.50	99	75-125						
Chlorobenzene	50.00	47.50	95	75-125						
1,1-Dichloroethene	50.00	52.00	104	75-125						
Methyl-tert-butyl ether (MTBE)	50.00	52.00	104	75-125						
Toluene (Methyl benzene)	50.00	47.00	94	75-125						
Trichloroethene	50.00	49.00	98	75-125						
LCS										
Chloroform (Trichloromethane)	50.00	68.00	136	75-125						
Ethylbenzene	50.00	45.50	91	75-125						
1,1,1-Trichloroethane	50.00	58.00	116	75-125						
o-Xylene	50.00	44.50	89	75-125						
m,p-Xylenes	100.00	88.00	88	75-125						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 18
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QUALITY CONTROL REPORT

QC Batch No: 072406 Sample Spiked: 072406 QC Prepared: 07/24/2006 QC Analyzed: 07/24/2006 Units: ug/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Benzene	0.0	50.00	51.50	103	50.00	49.00	98	5.0	75-125	<20
Chlorobenzene	0.0	50.00	48.00	96	50.00	44.50	89	7.6	75-125	<20
1,1-Dichloroethene	0.0	50.00	49.50	99	50.00	48.50	97	2.0	75-125	<20
Methyl-tert-butyl ether (MTBE)	0.0	50.00	59.50	119	50.00	55.50	111	7.0	75-125	<20
Toluene (Methyl benzene)	0.0	50.00	49.50	99	50.00	45.50	91	8.4	75-125	<20
Trichloroethene	0.0	50.00	51.50	103	50.00	48.50	97	6.0	75-125	<20

QC Batch No: 072406 Sample Spiked: 072406 QC Prepared: 07/24/2006 QC Analyzed: 07/24/2006 Units: ug/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Benzene	50.00	51.50	103	75-125						
Chlorobenzene	50.00	48.50	97	75-125						
1,1-Dichloroethene	50.00	47.50	95	75-125						
Methyl-tert-butyl ether (MTBE)	50.00	60.50	121	75-125						
Toluene (Methyl benzene)	50.00	50.00	100	75-125						
Trichloroethene	50.00	53.00	106	75-125						
LCS										
Chloroform (Trichloromethane)	50.00	56.00	112	75-125						
Ethylbenzene	50.00	48.50	97	75-125						
1,1,1-Trichloroethane	50.00	53.00	106	75-125						
o-Xylene	50.00	48.50	97	75-125						
m,p-Xylenes	100.00	95.00	95	75-125						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 19
Project ID: 32486-004
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38202	07/14/2006	H&A

Method: (8260B), Volatile Organic Compounds by GC/MS (SW846)

QUALITY CONTROL REPORT

QC Batch No: 072506 Sample Spiked: 072506 QC Prepared: 07/25/2006 QC Analyzed: 07/25/2006 Units: ug/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Benzene	0.0	50.00	49.50	99	50.00	48.00	96	3.1	75-125	<20
Chlorobenzene	0.0	50.00	44.00	88	50.00	42.00	84	4.7	75-125	<20
1,1-Dichloroethene	0.0	50.00	47.50	95	50.00	43.50	87	8.8	75-125	<20
Methyl-tert-butyl ether (MTBE)	0.0	50.00	59.50	119	50.00	61.00	122	2.5	75-125	<20
Toluene (Methyl benzene)	0.0	50.00	45.50	91	50.00	45.00	90	1.1	75-125	<20
Trichloroethene	0.0	50.00	49.00	98	50.00	47.00	94	4.2	75-125	<20

QC Batch No: 072506 Sample Spiked: 072506 QC Prepared: 07/25/2006 QC Analyzed: 07/25/2006 Units: ug/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Benzene	50.00	51.00	102	75-125						
Chlorobenzene	50.00	43.50	87	75-125						
1,1-Dichloroethene	50.00	46.00	92	75-125						
Methyl-tert-butyl ether (MTBE)	50.00	61.50	123	75-125						
Toluene (Methyl benzene)	50.00	48.00	96	75-125						
Trichloroethene	50.00	49.50	99	75-125						
LCS										
Chloroform (Trichloromethane)	50.00	61.00	122	75-125						
Ethylbenzene	50.00	44.00	88	75-125						
1,1,1-Trichloroethane	50.00	57.00	114	75-125						
o-Xylene	50.00	44.50	89	75-125						
m,p-Xylenes	100.00	82.00	82	75-125						

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

Data Qualifier:

*:	In the QC section, sample results have been taken directly from the ICP reading. No preparation factor has been applied.
B:	Analyte was present in the Method Blank.
D:	Result is from a diluted analysis.
E:	Result is beyond calibration limits and is estimated.
H:	Analysis was performed over the allowed holding time due to circumstances which were beyond laboratory control.
J:	Analyte was detected . However, the analyte concentration is an estimated value, which is between the Method Detection Limit (MDL) and the Practical Quantitation Limit (PQL).
M:	Matrix spike recovery is outside control limits due to matrix interference. Laboratory Control Sample recovery was acceptable.
S6:	Surrogate recovery is outside control limits due to matrix interference.
S8:	The analysis of the sample required a dilution such that the surrogate concentration was diluted below the method acceptance criteria.
X:	Results represent LCS and LCSD data.

Definition:

%Limi:	Percent acceptable limits.
%REC:	Percent recovery.
Con.L:	Acceptable Control Limits
Conce:	Added concentration to the sample.
LCS:	Laboratory Control Sample
MDL:	Method Detection Limit is a statistically derived number which is specific for each instrument, each method, and each compound. It indicates a distinctively detectable quantity with 99% probability.
MS:	Matrix Spike
MS DU:	Matrix Spike Duplicate

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

ND: Analyte was not detected in the sample at or above MDL.

PQL: Practical Quantitation Limit or ML (Minimum Level as per RWQCB) is the minimum concentration that can be quantified with more than 99% confidence. Taking into account all aspects of the entire analytical instrumentation and practice.

Recov: Recovered concentration in the sample.

RPD: Relative Percent Difference

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

RECEIVED
AUG 21 2006
Haley & Aldrich, Inc.

Ordered By

Haley & Aldrich
500 South Kraemer Blvd. Suite 370
Brea, CA 92821-6723

Telephone: (714) 985-3434
Attention: Tom Tatnall

Number of Pages 12
Date Received 08/07/2006
Date Reported 08/14/2006

Job Number	Order Date	Client
38422	08/07/2006	H&A

Project ID: 32486-007
Project Name: Delphi System Corp.
Site: 1201 N. Magnolia Avenue
Anaheim, CA 92801

Enclosed please find results of analyses of 12 soil samples which were analyzed as specified on the attached chain of custody. If there are any questions, please do not hesitate to call.

Checked By: _____ Approved By:

Cyrus Razmara, Ph.D.
Laboratory Director

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

CHAIN OF CUSTODY RECORD

№ 42880

AETL JOB No. # 38422

Page 1 of 10

COMPANY: Haley + Aldrich
PROJECT MANAGER: T. Tatnall
COMPANY ADDRESS: 500 S. Kraemer Blvd #370
PHONE: 714-985-3434
FAX: 714-985-3433
PROJECT NAME: Delphi - Anaheim
PROJECT #: 32486-007
SITE NAME AND ADDRESS: 1201 N. Magnolia, Anaheim
PO #:

ANALYSIS REQUESTED

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	PCBs 8082	Arsenic 6010B	Lead 6010B	Hold							TEST INSTRUCTIONS & COMMENTS
1	DP0276_SS_000_01		8-7-06	805	S	1 sleeve		X	X									38422-01
2	001_01			806				X	X									38422.02
3	002_01			807							X							38422.03
4	003_01			808							X							38422.04
5	004_01			809							X							38422.05
6	DP0274_SS_000_01			815				X	X									38422.06
7	001_01			816				X	X									38422.07
8	002_01			817							X							38422.08
9	003_01			818							X							38422.09
10	004_01			819							X							38422.10
11	DP0275_SS_000_01			825				X	X									38422.11
12	001_01			826				X	X									38422.12
13	002_01			827							X							38422.13
14	003_01			828							X							38422.14
15	004_01			829							X							38422.15

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y / N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☒ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☐ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: [signature]	Signature:	Signature: [signature]
Printed Name: R. Fry	Printed Name:	Printed Name: Thomas Robinson
Date: 8/7/06 Time: 1510	Date: Time:	Date: 08-07-06 Time: 1640
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: [signature]	Signature:	Signature: [signature]
Printed Name: Thomas Robinson	Printed Name:	Printed Name: A. Gisiki
Date: 08-07-06 Time: 1510	Date: Time:	Date: 8/7-06 Time: 1640

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

CHAIN OF CUSTODY RECORD
№ 42879

AETL JOB No. #38422

COMPANY: Haley + Aldrich
PROJECT MANAGER: T Tathall
COMPANY ADDRESS: 500 S. Kraemer Blvd #370
PHONE: 714-985-3434
FAX: 714-985-3433
PROJECT NAME: Delphi Anaheim
PROJECT #: 32486-0007
SITE NAME AND ADDRESS: 1201 N. Magnolia Anaheim
PO #:

	SAMPLE ID	LAB ID	DATE	TIME	MATRIX	CONTAINER NUMBER/SIZE	PRES.	PCBS 8082	Metals AS	Hold							TEST INSTRUCTIONS & COMMENTS
1	DP0273_SS_000_01		8-7-06	835	S	1 sleeve		X	X								38422.16
2	001_01			836				X	X								38422.17
3	002_01			837						X							38422.18
4	003_01			838						X							38422.19
5	004_01			839						X							38422.20
6	DP0272_SS_000_01			845				X	X								38422.21
7	001_01			846				X	X								38422.22
8	002_01			847						X							38422.23
9	003_01			848						X							38422.24
10	004_01			849						X							38422.25
11	DP0271_SS_000_01			853				X	X								38422.26
12	001_01			854				X	X								38422.27
13	002_01			855						X							38422.28
14	003_01			856						X							38422.29
15	004_01			857						X							38422.30

SAMPLE RECEIPT - TO BE FILLED BY LABORATORY

TOTAL NUMBER OF CONTAINERS		PROPERLY COOLED Y / N / NA	
CUSTODY SEALS Y / N / NA		SAMPLES INTACT Y/ N / NA	
RECEIVED IN GOOD COND. Y / N		SAMPLES ACCEPTED Y / N	

TURN AROUND TIME

☒ NORMAL ☐ RUSH ☐ SAME DAY ☐ 24 HRS. ☐ 48 HRS. ☐ 72 HRS.

RELINQUISHED BY SAMPLER: 1.	RELINQUISHED BY: 2.	RELINQUISHED BY: 3.
Signature: K. Hagan	Signature:	Signature: [illegible]
Printed Name: K. Hagan	Printed Name:	Printed Name: Thomas Robinson
Date: 8/7/06 Time: 1610	Date: Time:	Date: 08-07-06 Time: 1640
RECEIVED BY: 1.	**RECEIVED BY: 2.**	**RECEIVED BY LABORATORY: 3.**
Signature: [illegible]	Signature:	Signature: [illegible]
Printed Name: Thomas Robinson	Printed Name:	Printed Name: Aguilar
Date: 08-07-06 Time: 1610	Date: Time:	Date: 8/7-06 Time: 1640

DISTRIBUTION: WHITE - Laboratory, CANARY - Laboratory, PINK - Project/Account Manager, YELLOW - Sampler/Originator

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	TCLP Limit mg/L	TTLC Limit mg/Kg	Units	DP0271-SS-000-01	DP0271-SS-001-01	DP0272-SS-000-01	DP0272-SS-001-01	DP0273-SS-000-01	DP0273-SS-001-01	DP0274-SS-000-01	DP0274-SS-001-01	DP0275-SS-000-01
Sample Date					8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006
Laboratory Job Number					38422	38422	38422	38422	38422	38422	38422	38422	38422
Arsenic	5.0	5.0	500	mg/Kg	**11.3**	**2.05J**	**28.6**	**0.85J**	**2.00J**	**1.75J**	**8.85**	**1.45J**	**3.15J**

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: SO_6010BSCAN.xls

TABLE
Summary of Soil Analytical Results for Arsenic by ICP
1201 N. Magnolia Avenue Anaheim, CA 92801

Arsenic by ICP. Method: (6010BSCAN)

Sample ID	STLC Limit mg/L	DP0275-SS-001-01	DP0276-SS-000-01	DP0276-SS-001-01
Sample Date		8/7/2006	8/7/2006	8/7/2006
Laboratory Job Number		38422	38422	38422
Arsenic	5.0	**2.05J**	**4.15J**	**1.40J**

File: SO_6010BSCAN.xls

TABLE
Summary of Soil Analytical Results for Polychlorinated Biphenyls (PCBs) by GC
1201 N. Magnolia Avenue Anaheim, CA 92801

Polychlorinated Biphenyls (PCBs) by GC. Method: (8082)

Sample ID	Units	DP0271-SS-000-01	DP0271-SS-001-01	DP0272-SS-000-01	DP0272-SS-001-01	DP0273-SS-000-01	DP0273-SS-001-01	DP0274-SS-000-01	DP0274-SS-001-01	DP0275-SS-000-01	DP0275-SS-001-01	DP0276-SS-000-01	DP0276-SS-001-01
Sample Date		8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006	8/7/2006
Laboratory Job Number		38422	38422	38422	38422	38422	38422	38422	38422	38422	38422	38422	38422
Aroclor-1016 (PCB-1016)	ug/Kg	ND<1250	ND<25.0	ND<12500	ND<25.0	ND<25.0	ND<25.0	ND<5000	ND<100	ND<100	ND<25.0	ND<25.0	ND<25.0
Aroclor-1221 (PCB-1221)	ug/Kg	ND<1250	ND<25.0	ND<12500	ND<25.0	ND<25.0	ND<25.0	ND<5000	ND<100	ND<100	ND<25.0	ND<25.0	ND<25.0
Aroclor-1232 (PCB-1232)	ug/Kg	ND<1250	ND<25.0	ND<12500	ND<25.0	ND<25.0	ND<25.0	ND<5000	ND<100	ND<100	ND<25.0	ND<25.0	ND<25.0
Aroclor-1242 (PCB-1242)	ug/Kg	ND<1250	ND<25.0	ND<12500	ND<25.0	ND<25.0	ND<25.0	ND<5000	ND<100	ND<100	ND<25.0	ND<25.0	ND<25.0
Aroclor-1248 (PCB-1248)	ug/Kg	**15,700**	**29.3J**	**214,000**	**40.5J**	**367**	ND<25.0	**139,000**	**2,330**	**1,850**	**1,060**	ND<25.0	ND<25.0
Aroclor-1254 (PCB-1254)	ug/Kg	ND<1250	ND<25.0	ND<12500	ND<25.0	ND<25.0	ND<25.0	ND<5000	ND<100	ND<100	ND<25.0	ND<25.0	ND<25.0
Aroclor-1260 (PCB-1260)	ug/Kg	ND<1250	ND<25.0	ND<12500	ND<25.0	ND<25.0	ND<25.0	ND<5000	ND<100	ND<100	ND<25.0	ND<25.0	ND<25.0
Aroclor-1262 (PCB-1262)	ug/Kg	ND<1250	ND<25.0	ND<12500	ND<25.0	ND<25.0	ND<25.0	ND<5000	ND<100	ND<100	ND<25.0	ND<25.0	ND<25.0
Aroclor-1268 (PCB-1268)	ug/Kg	ND<1250	ND<25.0	ND<12500	ND<25.0	ND<25.0	ND<25.0	ND<5000	ND<100	ND<100	ND<25.0	ND<25.0	ND<25.0

1) "ND<X" INDICATES CONSTITUENT(S) NOT DETECTED AT OR ABOVE METHOD DETECTION LIMIT.
2) "J" INDICATES ANALYTE WAS DETECTED. HOWEVER, ANALYTE CONCENTRATION IS AN ESTIMATED VALUE WHICH IS BETWEEN THE METHOD DETECTION LIMIT (MDL) AND THE PRACTICAL QUANTITATION LIMIT(PQL).
3) "D" INDICATES THE SAMPLE WAS DILUTED TO BRING THE ANALYTE CONCENTRATION WITHIN CALIBRATION RANGE.

File: SO_8082.xls

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 2
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 080806

Our Lab I.D.			Method Blank	38422.01	38422.02		
Client Sample I.D.				DP0276-SS-000-01	DP0276-SS-001-01		
Date Sampled				08/07/2006	08/07/2006		
Date Prepared			08/08/2006	08/08/2006	08/08/2006		
Preparation Method			3550B	3550B	3550B		
Date Analyzed			08/08/2006	08/08/2006	08/08/2006		
Matrix			Soil	Soil	Soil		
Units			ug/Kg	ug/Kg	ug/Kg		
Dilution Factor			1	1	1		
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**		
Aroclor-1016 (PCB-1016)	25.0	50.0	ND	ND	ND		
Aroclor-1221 (PCB-1221)	25.0	50.0	ND	ND	ND		
Aroclor-1232 (PCB-1232)	25.0	50.0	ND	ND	ND		
Aroclor-1242 (PCB-1242)	25.0	50.0	ND	ND	ND		
Aroclor-1248 (PCB-1248)	25.0	50.0	ND	ND	ND		
Aroclor-1254 (PCB-1254)	25.0	50.0	ND	ND	ND		
Aroclor-1260 (PCB-1260)	25.0	50.0	ND	ND	ND		
Aroclor-1262 (PCB-1262)	25.0	50.0	ND	ND	ND		
Aroclor-1268 (PCB-1268)	25.0	50.0	ND	ND	ND		

Our Lab I.D.				38422.01	38422.02		
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**		
Tetrachloro-m-xylene	30-150		93	114	117		

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 3
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 080806

Our Lab I.D.			38422.06
Client Sample I.D.			DP0274-SS-0 00-01
Date Sampled			08/07/2006
Date Prepared			08/08/2006
Preparation Method			3550B
Date Analyzed			08/10/2006
Matrix			Soil
Units			ug/Kg
Dilution Factor			200
Analytes	**MDL**	**PQL**	**Results**
Aroclor-1016 (PCB-1016)	5000	10000	ND
Aroclor-1221 (PCB-1221)	5000	10000	ND
Aroclor-1232 (PCB-1232)	5000	10000	ND
Aroclor-1242 (PCB-1242)	5000	10000	ND
Aroclor-1248 (PCB-1248)	5000	10000	139,000
Aroclor-1254 (PCB-1254)	5000	10000	ND
Aroclor-1260 (PCB-1260)	5000	10000	ND
Aroclor-1262 (PCB-1262)	5000	10000	ND
Aroclor-1268 (PCB-1268)	5000	10000	ND

Our Lab I.D.			38422.06
Surrogates	**%Rec.Limit**		**% Rec.**
Tetrachloro-m-xylene	30-150		0 S8

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **4**
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 080806

Our Lab I.D.			**38422.07**	**38422.11**			
Client Sample I.D.			DP0274-SS-001-01	DP0275-SS-000-01			
Date Sampled			08/07/2006	08/07/2006			
Date Prepared			08/08/2006	08/08/2006			
Preparation Method			3550B	3550B			
Date Analyzed			08/09/2006	08/09/2006			
Matrix			Soil	Soil			
Units			ug/Kg	ug/Kg			
Dilution Factor			4	4			
Analytes	**MDL**	**PQL**	**Results**	**Results**			
Aroclor-1016 (PCB-1016)	100	200	ND	ND			
Aroclor-1221 (PCB-1221)	100	200	ND	ND			
Aroclor-1232 (PCB-1232)	100	200	ND	ND			
Aroclor-1242 (PCB-1242)	100	200	ND	ND			
Aroclor-1248 (PCB-1248)	100	200	2,330	1,850			
Aroclor-1254 (PCB-1254)	100	200	ND	ND			
Aroclor-1260 (PCB-1260)	100	200	ND	ND			
Aroclor-1262 (PCB-1262)	100	200	ND	ND			
Aroclor-1268 (PCB-1268)	100	200	ND	ND			

Our Lab I.D.			**38422.07**	**38422.11**			
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**			
Tetrachloro-m-xylene	30-150		114	93			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 5
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 080806

Our Lab I.D.			38422.12	38422.16	38422.17		
Client Sample I.D.			DP0275-SS-0 01-01	DP0273-SS-0 00-01	DP0273-SS-0 01-01		
Date Sampled			08/07/2006	08/07/2006	08/07/2006		
Date Prepared			08/08/2006	08/08/2006	08/08/2006		
Preparation Method			3550B	3550B	3550B		
Date Analyzed			08/08/2006	08/08/2006	08/08/2006		
Matrix			Soil	Soil	Soil		
Units			ug/Kg	ug/Kg	ug/Kg		
Dilution Factor			1	1	1		
Analytes	**MDL**	**PQL**	**Results**	**Results**	**Results**		
Aroclor-1016 (PCB-1016)	25.0	50.0	ND	ND	ND		
Aroclor-1221 (PCB-1221)	25.0	50.0	ND	ND	ND		
Aroclor-1232 (PCB-1232)	25.0	50.0	ND	ND	ND		
Aroclor-1242 (PCB-1242)	25.0	50.0	ND	ND	ND		
Aroclor-1248 (PCB-1248)	25.0	50.0	1,060	367	ND		
Aroclor-1254 (PCB-1254)	25.0	50.0	ND	ND	ND		
Aroclor-1260 (PCB-1260)	25.0	50.0	ND	ND	ND		
Aroclor-1262 (PCB-1262)	25.0	50.0	ND	ND	ND		
Aroclor-1268 (PCB-1268)	25.0	50.0	ND	ND	ND		

Our Lab I.D.			38422.12	38422.16	38422.17		
Surrogates	**%Rec.Limit**		**% Rec.**	**% Rec.**	**% Rec.**		
Tetrachloro-m-xylene	30-150		107	105	109		

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 6
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 080806

Our Lab I.D.			**38422.21**
Client Sample I.D.			DP0272-SS-0 00-01
Date Sampled			08/07/2006
Date Prepared			08/08/2006
Preparation Method			3550B
Date Analyzed			08/10/2006
Matrix			Soil
Units			ug/Kg
Dilution Factor			500
Analytes	**MDL**	**PQL**	**Results**
Aroclor-1016 (PCB-1016)	12500	25000	ND
Aroclor-1221 (PCB-1221)	12500	25000	ND
Aroclor-1232 (PCB-1232)	12500	25000	ND
Aroclor-1242 (PCB-1242)	12500	25000	ND
Aroclor-1248 (PCB-1248)	12500	25000	214,000
Aroclor-1254 (PCB-1254)	12500	25000	ND
Aroclor-1260 (PCB-1260)	12500	25000	ND
Aroclor-1262 (PCB-1262)	12500	25000	ND
Aroclor-1268 (PCB-1268)	12500	25000	ND

Our Lab I.D.			**38422.21**
Surrogates	**%Rec.Limit**		**% Rec.**
Tetrachloro-m-xylene	30-150		0 S8

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 7
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 080806

Our Lab I.D.			**38422.22**
Client Sample I.D.			DP0272-SS-0 01-01
Date Sampled			08/07/2006
Date Prepared			08/08/2006
Preparation Method			3550B
Date Analyzed			08/09/2006
Matrix			Soil
Units			ug/Kg
Dilution Factor			1
Analytes	**MDL**	**PQL**	**Results**
Aroclor-1016 (PCB-1016)	25.0	50.0	ND
Aroclor-1221 (PCB-1221)	25.0	50.0	ND
Aroclor-1232 (PCB-1232)	25.0	50.0	ND
Aroclor-1242 (PCB-1242)	25.0	50.0	ND
Aroclor-1248 (PCB-1248)	25.0	50.0	40.5J
Aroclor-1254 (PCB-1254)	25.0	50.0	ND
Aroclor-1260 (PCB-1260)	25.0	50.0	ND
Aroclor-1262 (PCB-1262)	25.0	50.0	ND
Aroclor-1268 (PCB-1268)	25.0	50.0	ND

Our Lab I.D.		**38422.22**
Surrogates	**%Rec.Limit**	**% Rec.**
Tetrachloro-m-xylene	30-150	95

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 8
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 080806

Our Lab I.D.			**38422.26**
Client Sample I.D.			DP0271-SS-0 00-01
Date Sampled			08/07/2006
Date Prepared			08/08/2006
Preparation Method			3550B
Date Analyzed			08/09/2006
Matrix			Soil
Units			ug/Kg
Dilution Factor			50
Analytes	**MDL**	**PQL**	**Results**
Aroclor-1016 (PCB-1016)	1250	2500	ND
Aroclor-1221 (PCB-1221)	1250	2500	ND
Aroclor-1232 (PCB-1232)	1250	2500	ND
Aroclor-1242 (PCB-1242)	1250	2500	ND
Aroclor-1248 (PCB-1248)	1250	2500	15,700
Aroclor-1254 (PCB-1254)	1250	2500	ND
Aroclor-1260 (PCB-1260)	1250	2500	ND
Aroclor-1262 (PCB-1262)	1250	2500	ND
Aroclor-1268 (PCB-1268)	1250	2500	ND

Our Lab I.D.		**38422.26**
Surrogates	**%Rec.Limit**	**% Rec.**
Tetrachloro-m-xylene	30-150	0 S8

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: 9
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QC Batch No: 080806

Our Lab I.D.			**38422.27**
Client Sample I.D.			DP0271-SS-0 01-01
Date Sampled			08/07/2006
Date Prepared			08/08/2006
Preparation Method			3550B
Date Analyzed			08/08/2006
Matrix			Soil
Units			ug/Kg
Dilution Factor			1
Analytes	**MDL**	**PQL**	**Results**
Aroclor-1016 (PCB-1016)	25.0	50.0	ND
Aroclor-1221 (PCB-1221)	25.0	50.0	ND
Aroclor-1232 (PCB-1232)	25.0	50.0	ND
Aroclor-1242 (PCB-1242)	25.0	50.0	ND
Aroclor-1248 (PCB-1248)	25.0	50.0	29.3J
Aroclor-1254 (PCB-1254)	25.0	50.0	ND
Aroclor-1260 (PCB-1260)	25.0	50.0	ND
Aroclor-1262 (PCB-1262)	25.0	50.0	ND
Aroclor-1268 (PCB-1268)	25.0	50.0	ND

Our Lab I.D.		**38422.27**
Surrogates	**%Rec.Limit**	**% Rec.**
Tetrachloro-m-xylene	30-150	108

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714) 985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page 10
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Analytes			Arsenic			
Methods of Analyses			(6010BSCAN)			
Date Prepared			08/10/2006			
Date Analyzed			08/10/2006			
Matrix			Soil			
QC Batch Number			081006			
Units			mg/Kg			
Method Detection Limit			0.5			
Practical Quantitation Limit			5.0			
Dilution Factor			1			
Lab ID	**Sample ID**	**Sampled**	**Results**			
38422.01	DP0276-SS-000-01	08/07/2006	4.15J			
38422.02	DP0276-SS-001-01	08/07/2006	1.40J			
38422.06	DP0274-SS-000-01	08/07/2006	8.85			
38422.07	DP0274-SS-001-01	08/07/2006	1.45J			
38422.11	DP0275-SS-000-01	08/07/2006	3.15J			
38422.12	DP0275-SS-001-01	08/07/2006	2.05J			
38422.16	DP0273-SS-000-01	08/07/2006	2.00J			
38422.17	DP0273-SS-001-01	08/07/2006	1.75J			
38422.21	DP0272-SS-000-01	08/07/2006	28.6			
38422.22	DP0272-SS-001-01	08/07/2006	0.85J			
38422.26	DP0271-SS-000-01	08/07/2006	11.3			
38422.27	DP0271-SS-001-01	08/07/2006	2.05J			
N/A	Method Blank	/ /	ND			

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **11**
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (6010BSCAN), Arsenic by ICP

QUALITY CONTROL REPORT

QC Batch No: 081006 Sample Spiked: 38422.02 QC Prepared: 08/10/2006 QC Analyzed: 08/10/2006 Units: mg/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Arsenic	0.028	1.00	0.96	93	1.00	0.96	93	<1	80-120	<15

QC Batch No: 081006 Sample Spiked: 38422.02 QC Prepared: 08/10/2006 QC Analyzed: 08/10/2006 Units: mg/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Arsenic	1.00	0.90	90	80-120						

American Environmental Testing Laboratory Inc.

2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

ANALYTICAL RESULTS

Ordered By

Haley & Aldrich
500 South Kraemer Blvd.
Suite 370
Brea, CA 92821-6723

Telephone: (714)985-3434
Attn: Tom Tatnall

Site

1201 N. Magnolia Avenue
Anaheim, CA 92801

Page: **12**
Project ID: 32486-007
Project Name: Delphi System Corp.

AETL Job Number	Submitted	Client
38422	08/07/2006	H&A

Method: (8082), Polychlorinated Biphenyls (PCBs) by GC

QUALITY CONTROL REPORT

QC Batch No: 080806 Sample Spiked: 080806 QC Prepared: 08/08/2006 QC Analyzed: 08/08/2006 Units: ug/Kg

Analytes	Sample Result	MS Concen	MS Recov	MS % REC	MS DUP Concen	MS DUP Recov	MS DUP % REC	RPD %	MS/MSD % Limit	MS RPD % Limit
Aroclor-1260 (PCB-1260)	0.0	500.00	450.00	90	500.00	440.00	88	2.2	50-150	<50

QC Batch No: 080806 Sample Spiked: 080806 QC Prepared: 08/08/2006 QC Analyzed: 08/08/2006 Units: ug/Kg

Analytes	LCS Concen	LCS Recov	LCS % REC	LCS/LCSD % Limit						
Aroclor-1260 (PCB-1260)	500.00	470.00	94	50-150						

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

Data Qualifier:

*:	In the QC section, sample results have been taken directly from the ICP reading. No preparation factor has been applied.
B:	Analyte was present in the Method Blank.
D:	Result is from a diluted analysis.
E:	Result is beyond calibration limits and is estimated.
H:	Analysis was performed over the allowed holding time due to circumstances which were beyond laboratory control.
J:	Analyte was detected . However, the analyte concentration is an estimated value, which is between the Method Detection Limit (MDL) and the Practical Quantitation Limit (PQL).
M:	Matrix spike recovery is outside control limits due to matrix interference. Laboratory Control Sample recovery was acceptable.
S6:	Surrogate recovery is outside control limits due to matrix interference.
S8:	The analysis of the sample required a dilution such that the surrogate concentration was diluted below the method acceptance criteria.
X:	Results represent LCS and LCSD data.

Definition:

%Limi:	Percent acceptable limits.
%REC:	Percent recovery.
Con.L:	Acceptable Control Limits
Conce:	Added concentration to the sample.
LCS:	Laboratory Control Sample
MDL:	Method Detection Limit is a statistically derived number which is specific for each instrument, each method, and each compound. It indicates a distinctively detectable quantity with 99% probability.
MS:	Matrix Spike
MS DU:	Matrix Spike Duplicate

American Environmental Testing Laboratory Inc.
2834 & 2908 North Naomi Street, Burbank, CA 91504 • DOHS NO: 1541, LACSD NO: 10181
Tel: (888) 288-AETL • (818) 845-8200 • Fax: (818) 845-8840 • www.aetlab.com

Data Qualifiers and Descriptors

ND: Analyte was not detected in the sample at or above MDL.

PQL: Practical Quantitation Limit or ML (Minimum Level as per RWQCB) is the minimum concentration that can be quantified with more than 99% confidence. Taking into account all aspects of the entire analytical instrumentation and practice.

Recov: Recovered concentration in the sample.

RPD: Relative Percent Difference